

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032123\
 Data File : PQ060579.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Mar 2023 16:46
 Operator : YP\AJ
 Sample : AR1254ICC800
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12544049

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 17:26:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 21 17:25:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.406	2.766	329.3E6	165.6E6	36.629	37.643
2) SA Decachlor...	8.584	7.547	408.7E6	327.2E6	73.034	74.941
Target Compounds						
26) L6 AR-1254-1	5.329	4.529	236.2E6	170.5E6	713.172	723.092
27) L6 AR-1254-2	5.546	4.676	355.5E6	147.5E6	720.083	729.621
28) L6 AR-1254-3	5.899	5.058	381.1E6	231.9E6	726.104	750.068
29) L6 AR-1254-4	6.184	5.284	275.0E6	146.7E6	713.543	745.452
30) L6 AR-1254-5	6.597	5.691	302.1E6	213.5E6	724.478	755.494

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032123\
Data File : PQ060579.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2023 16:46
Operator : YP\AJ
Sample : AR1254ICC800
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12544049

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 17:26:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 21 17:25:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

