

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032322\
 Data File : PQ056777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Mar 2022 03:20
 Operator : YP\AJ
 Sample : AR1262ICC1600
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12625045

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 08:15:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 08:06:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.531	3.748	1644.1E6	990.7E6	78.069	80.161
2) SA Decachlor...	10.300	8.686	2305.0E6	1740.1E6	152.334	159.918
Target Compounds						
36) L8 AR-1262-1	7.912	6.780	1471.6E6	668.4E6	1538.035	1603.127
37) L8 AR-1262-2	8.471	7.313	2814.3E6	2556.7E6	1581.438	1610.139
38) L8 AR-1262-3	8.783	7.593	1195.8E6	955.8E6	1557.359	1557.908
39) L8 AR-1262-4	8.877	7.657	908.2E6	1863.1E6	1563.688m	1604.071
40) L8 AR-1262-5	9.549	8.148	995.1E6	805.3E6	1550.081	1617.089

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032322\
Data File : PQ056777.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2022 03:20
Operator : YP\AJ
Sample : AR1262ICC1600
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12625045

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 08:15:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 08:06:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

