

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032323\
 Data File : PQ060645.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2023 11:40
 Operator : YP\AJ
 Sample : 01990-18
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 E0274

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/24/2023
 Supervised By :Ankita Jodhani 03/24/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 01:23:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 21 19:46:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.406	2.764	182.0E6	220.5E6	20.633	53.814 #
2) SA Decachlor...	8.584	7.544	160.9E6	167.3E6	29.121m	39.237 #
Target Compounds						
21) L5 AR-1248-1	4.506	3.769	205.8E6	108.2E6	1061.293	1190.046
22) L5 AR-1248-2	4.772	3.995	217.5E6	140.9E6	833.183	1038.044
23) L5 AR-1248-3	4.961	4.031	265.5E6	123.2E6	844.465	938.078
24) L5 AR-1248-4	5.355	4.189	235.9E6	159.0E6	674.400	988.021 #
25) L5 AR-1248-5	5.389	4.564	257.9E6	147.0E6	765.023	945.393

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032323\
Data File : PQ060645.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2023 11:40
Operator : YP\AJ
Sample : 01990-18
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
E0274

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/24/2023
Supervised By :Ankita Jodhani 03/24/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 01:23:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 21 19:46:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

