

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032419\
 Data File : PQ038214.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2019 18:13
 Operator : SM\SJ
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 04:18:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 04:17:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.729	4.016	200.5E6	85953296	45.772	46.032
2) SA Decachlor...	10.692	9.177	458.7E6	201.3E6	74.188	76.475
Target Compounds						
16) L4 AR-1242-1	5.912	5.128	124.7E6	58631003	869.388	909.522
17) L4 AR-1242-2	5.935	5.148	190.1E6	84703204	876.114	913.930
18) L4 AR-1242-3	5.999	5.329	117.6E6	44920222	865.536	906.155
19) L4 AR-1242-4	6.099	5.411	96431560	44613898	874.507	881.653
20) L4 AR-1242-5	6.838	5.940	112.5E6	62569901	856.537	892.662



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032419\
Data File : PQ038214.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2019 18:13
Operator : SM\SJ
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 04:18:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 04:17:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#16 AR-1242-1


R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 124718833
Conc: 869.39 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#16 AR-1242-1

R.T.: 5.128 min
Delta R.T.: 0.000 min
Response: 58631003
Conc: 909.52 ng/ml




#17 AR-1242-2

R.T.: 5.935 min
Delta R.T.: 0.000 min
Response: 190085368
Conc: 876.11 ng/ml



#17 AR-1242-2


R.T.: 5.148 min
Delta R.T.: 0.000 min
Response: 84703204
Conc: 913.93 ng/ml



#18 AR-1242-3


R.T.: 5.999 min
Delta R.T.: 0.000 min
Response: 117557563
Conc: 865.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#18 AR-1242-3

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 44920222
Conc: 906.16 ng/ml




#19 AR-1242-4

R.T.: 6.099 min
Delta R.T.: 0.000 min
Response: 96431560
Conc: 874.51 ng/ml



#19 AR-1242-4


R.T.: 5.411 min
Delta R.T.: 0.000 min
Response: 44613898
Conc: 881.65 ng/ml



#20 AR-1242-5

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 112467848
Conc: 856.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#20 AR-1242-5

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 62569901
Conc: 892.66 ng/ml