

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
 Data File : PQ060678.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2023 13:15
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 14:35:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 27 14:34:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.406	2.764	573.9E6	304.5E6	71.622	72.981
2)	SA Decachlor...	8.584	7.544	402.1E6	347.7E6	70.452	71.551
Target Compounds							
16)	L4 AR-1242-1	4.506	3.769	154.5E6	81765174	706.328	716.207
17)	L4 AR-1242-2	4.525	3.785	229.0E6	118.9E6	706.837	715.055
18)	L4 AR-1242-3	4.583	3.946	144.2E6	63387657	697.652	714.471
19)	L4 AR-1242-4	4.675	4.031	124.8E6	60346032	701.812	708.033
20)	L4 AR-1242-5	5.390	4.527	125.1E6	83558422	702.573	714.986



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
Data File : PQ060678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2023 13:15
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:35:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 27 14:34:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 573889088
Conc: 71.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.764 min
Delta R.T.: 0.000 min
Response: 304489295
Conc: 72.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: 0.000 min
Response: 402134509
Conc: 70.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.544 min
Delta R.T.: 0.000 min
Response: 347691501
Conc: 71.55 ng/ml



#16 AR-1242-1


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 154452713
Conc: 706.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#16 AR-1242-1

R.T.: 3.769 min
Delta R.T.: 0.000 min
Response: 81765174
Conc: 716.21 ng/ml




#17 AR-1242-2

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 229049197
Conc: 706.84 ng/ml



#17 AR-1242-2


R.T.: 3.785 min
Delta R.T.: 0.000 min
Response: 118877716
Conc: 715.06 ng/ml



#18 AR-1242-3


R.T.: 4.583 min
Delta R.T.: 0.000 min
Response: 144242278
Conc: 697.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#18 AR-1242-3

R.T.: 3.946 min
Delta R.T.: 0.000 min
Response: 63387657
Conc: 714.47 ng/ml




#19 AR-1242-4

R.T.: 4.675 min
Delta R.T.: 0.000 min
Response: 124766709
Conc: 701.81 ng/ml



#19 AR-1242-4


R.T.: 4.031 min
Delta R.T.: 0.000 min
Response: 60346032
Conc: 708.03 ng/ml



#20 AR-1242-5

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 125053152
Conc: 702.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#20 AR-1242-5

R.T.: 4.527 min
Delta R.T.: 0.000 min
Response: 83558422
Conc: 714.99 ng/ml