

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
 Data File : PQ060671.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2023 11:33
 Operator : YP\AJ
 Sample : AR1221ICC1000
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1221ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 14:54:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 27 14:54:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.407	2.765	764.5E6	401.2E6	95.017	98.021
2)	SA Decachlor...	8.585	7.545	527.8E6	458.9E6	94.453	97.039
Target Compounds							
8)	L2 AR-1221-1	3.606	2.968	93776231	46854548	911.504	935.137
9)	L2 AR-1221-2	3.685	3.044	69626402	35412000	916.847	943.136
10)	L2 AR-1221-3	3.755	3.112	203.0E6	107.2E6	905.052	927.087

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
Data File : PQ060671.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2023 11:33
Operator : YP\AJ
Sample : AR1221ICC1000
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 14:54:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 27 14:54:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

