

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
 Data File : PQ060697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2023 17:53
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ032723

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 18:08:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 27 17:20:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.406	2.763	387.6E6	206.8E6	49.372	52.602
2) SA Decachlor...	8.584	7.543	278.9E6	236.5E6	49.430	50.191
Target Compounds						
21) L5 AR-1248-1	4.506	3.768	85027343	44012664	503.641	498.981
22) L5 AR-1248-2	4.772	3.994	114.5E6	64837878	500.448	497.346
23) L5 AR-1248-3	4.962	4.030	139.2E6	62926032	496.796	501.227
24) L5 AR-1248-4	5.355	4.188	155.7E6	78150471	500.212	501.629
25) L5 AR-1248-5	5.390	4.564	151.7E6	76650078	506.112	499.692



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032723\
Data File : PQ060697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2023 17:53
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ032723

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 18:08:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 27 17:20:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 387577039
Conc: 49.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ032723




#1 Tetrachloro-m-xylene

R.T.: 2.763 min
Delta R.T.: 0.000 min
Response: 206757060
Conc: 52.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: 0.000 min
Response: 278936578
Conc: 49.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.543 min
Delta R.T.: -0.001 min
Response: 236548619
Conc: 50.19 ng/ml



#21 AR-1248-1


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 85027343
Conc: 503.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ032723




#21 AR-1248-1

R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 44012664
Conc: 498.98 ng/ml




#22 AR-1248-2

R.T.: 4.772 min
Delta R.T.: -0.001 min
Response: 114547192
Conc: 500.45 ng/ml



#22 AR-1248-2


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 64837878
Conc: 497.35 ng/ml



#23 AR-1248-3


R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 139208776
Conc: 496.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ032723




#23 AR-1248-3

R.T.: 4.030 min
Delta R.T.: -0.001 min
Response: 62926032
Conc: 501.23 ng/ml




#24 AR-1248-4

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 155723182
Conc: 500.21 ng/ml



#24 AR-1248-4


R.T.: 4.188 min
Delta R.T.: -0.001 min
Response: 78150471
Conc: 501.63 ng/ml



#25 AR-1248-5

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 151740927
Conc: 506.11 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ032723



#25 AR-1248-5

R.T.: 4.564 min
Delta R.T.: 0.000 min
Response: 76650078
Conc: 499.69 ng/ml