

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033123\
 Data File : PQ060732.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2023 17:08
 Operator : YP\AJ
 Sample : 02152-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 EYG07

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/03/2023
 Supervised By :Ankita Jodhani 04/03/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 04:53:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 21 19:46:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.405	2.763	156.2E6	77524217	17.706	18.922
2)	SA Decachlor...	8.584	7.543	174.3E6	147.3E6	31.542	34.539
Target Compounds							
26)	L6 AR-1254-1	5.326	4.525	19510451	20080455	57.938	85.284m#
27)	L6 AR-1254-2	5.540	4.671	21655174	8808359	43.066	43.262m
28)	L6 AR-1254-3	5.899	5.053	31499601	15542966	58.744	49.467m
29)	L6 AR-1254-4	6.183	5.280	25385004	15891953	65.446	79.586m
30)	L6 AR-1254-5	6.596	5.689	19477079	31463437	45.911	108.651m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033123\
Data File : PQ060732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Mar 2023 17:08
Operator : YP\AJ
Sample : 02152-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
EYG07

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/03/2023
Supervised By :Ankita Jodhani 04/03/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 04:53:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 21 19:46:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

