

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040221\
 Data File : PR049704.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2021 13:38
 Operator : DD\AJ
 Sample : PR040221ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 08:14:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 03 08:12:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	3.904	3.911 E6	-0.2	95	0.00
2 SA Decachlorobiphenyl	4.443	4.488 E6	-1.0	101	0.00
3 L1 AR-1016-1	153.726	145.329 E3	5.5	94	0.00
4 L1 AR-1016-2	213.649	206.114 E3	3.5	95	0.00
5 L1 AR-1016-3	128.399	123.353 E3	3.9	95	0.00
6 L1 AR-1016-4	106.407	103.098 E3	3.1	95	0.00
7 L1 AR-1016-5	107.140	104.113 E3	2.8	95	0.00
31 L7 AR-1260-1	198.861	196.303 E3	1.3	99	0.00
32 L7 AR-1260-2	244.130	242.068 E3	0.8	99	0.00
33 L7 AR-1260-3	186.968	186.226 E3	0.4	99	0.00
34 L7 AR-1260-4	218.486	219.200 E3	-0.3	100	0.00
35 L7 AR-1260-5	451.075	458.918 E3	-1.7	100	0.00

Signal #2

1 SA Tetrachloro-m-xylene	6.632	6.743 E6	-1.7	97	0.00
2 SA Decachlorobiphenyl	7.613	7.717 E6	-1.4	98	0.00
3 L1 AR-1016-1	240.120	235.775 E3	1.8	95	0.00
4 L1 AR-1016-2	344.975	339.339 E3	1.6	95	0.00
5 L1 AR-1016-3	182.063	178.659 E3	1.9	96	0.00
6 L1 AR-1016-4	139.028	134.058 E3	3.6	96	0.00
7 L1 AR-1016-5	187.885	181.831 E3	3.2	96	0.00
31 L7 AR-1260-1	358.659	354.025 E3	1.3	97	0.00
32 L7 AR-1260-2	443.966	439.531 E3	1.0	97	0.00
33 L7 AR-1260-3	420.647	420.486 E3	0.0	97	0.00
34 L7 AR-1260-4	356.400	360.493 E3	-1.1	98	0.00
35 L7 AR-1260-5	916.305	937.213 E3	-2.3	98	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	41.336	0.000 E3	100.0#	0#	-4.85#
9 L2 AR-1221-2	31.200	0.000 E3	100.0#	0#	-4.94#
10 L2 AR-1221-3	94.911	0.000 E3	100.0#	0#	-5.02#
11 L3 AR-1232-1	77.271	0.000 E3	100.0#	0#	-5.02#
12 L3 AR-1232-2	42.382	0.000 E3	100.0#	0#	-5.55#
13 L3 AR-1232-3	85.196	0.000 E3	100.0#	0#	-5.84#
14 L3 AR-1232-4	42.620	0.000 E3	100.0#	0#	-6.01#
15 L3 AR-1232-5	33.070	0.000 E3	100.0#	0#	-6.09#
16 L4 AR-1242-1	112.120	0.000 E3	100.0#	0#	-5.82#
17 L4 AR-1242-2	157.959	0.000 E3	100.0#	0#	-5.84#
18 L4 AR-1242-3	95.765	0.000 E3	100.0#	0#	-5.90#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040221\
 Data File : PR049704.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2021 13:38
 Operator : DD\AJ
 Sample : PR040221ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 08:14:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 03 08:12:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	80.127	0.000 E3	100.0#	0#	-6.00#
20 L4 AR-1242-5	91.107	0.000 E3	100.0#	0#	-6.74#
21 L5 AR-1248-1	85.682	0.000 E3	100.0#	0#	-5.82#
22 L5 AR-1248-2	120.212	0.000 E3	100.0#	0#	-6.09#
23 L5 AR-1248-3	135.434	0.000 E3	100.0#	0#	-6.30#
24 L5 AR-1248-4	160.240	0.000 E3	100.0#	0#	-6.70#
25 L5 AR-1248-5	153.810	0.000 E3	100.0#	0#	-6.74#
26 L6 AR-1254-1	159.081	0.000 E3	100.0#	0#	-6.68#
27 L6 AR-1254-2	251.876	0.000 E3	100.0#	0#	-6.89#
28 L6 AR-1254-3	276.261	0.000 E3	100.0#	0#	-7.26#
29 L6 AR-1254-4	210.427	0.000 E3	100.0#	0#	-7.55#
30 L6 AR-1254-5	226.785	0.000 E3	100.0#	0#	-7.97#
36 L8 AR-1262-1	261.848	0.000 E3	100.0#	0#	-8.04#
37 L8 AR-1262-2	497.652	0.000 E3	100.0#	0#	-8.61#
38 L8 AR-1262-3	328.883	0.000 E3	100.0#	0#	-8.94#
39 L8 AR-1262-4	247.032	0.000 E3	100.0#	0#	-9.03#
40 L8 AR-1262-5	180.195	0.000 E3	100.0#	0#	-9.73#
41 L9 AR-1268-1	586.487	0.000 E3	100.0#	0#	-8.94#
42 L9 AR-1268-2	538.386	0.000 E3	100.0#	0#	-9.04#
43 L9 AR-1268-3	457.599	0.000 E3	100.0#	0#	-9.27#
44 L9 AR-1268-4	201.961	0.000 E3	100.0#	0#	-9.73#
45 L9 AR-1268-5	1.636	0.000 E6	100.0#	0#	-10.16#

Signal #2

8 L2 AR-1221-1	63.729	0.000 E3	100.0#	0#	-4.04#
9 L2 AR-1221-2	49.147	0.000 E3	100.0#	0#	-4.12#
10 L2 AR-1221-3	157.384	0.000 E3	100.0#	0#	-4.20#
11 L3 AR-1232-1	126.072	0.000 E3	100.0#	0#	-4.20#
12 L3 AR-1232-2	132.829	0.000 E3	100.0#	0#	-4.93#
13 L3 AR-1232-3	69.528	0.000 E3	100.0#	0#	-5.11#
14 L3 AR-1232-4	58.788	0.000 E3	100.0#	0#	-5.19#
15 L3 AR-1232-5	66.018	0.000 E3	100.0#	0#	-5.36#
16 L4 AR-1242-1	175.497	0.000 E3	100.0#	0#	-4.91#
17 L4 AR-1242-2	251.055	0.000 E3	100.0#	0#	-4.93#
18 L4 AR-1242-3	133.410	0.000 E3	100.0#	0#	-5.11#
19 L4 AR-1242-4	123.935	0.000 E3	100.0#	0#	-5.19#
20 L4 AR-1242-5	170.646	0.000 E3	100.0#	0#	-5.71#
21 L5 AR-1248-1	134.949	0.000 E3	100.0#	0#	-4.91#
22 L5 AR-1248-2	177.225	0.000 E3	100.0#	0#	-5.15#
23 L5 AR-1248-3	188.437	0.000 E3	100.0#	0#	-5.19#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040221\
 Data File : PR049704.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2021 13:38
 Operator : DD\AJ
 Sample : PR040221ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 08:14:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 03 08:12:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	232.643	0.000 E3	100.0#	0#	-5.36#
25 L5 AR-1248-5	246.302	0.000 E3	100.0#	0#	-5.75#
26 L6 AR-1254-1	375.931	0.000 E3	100.0#	0#	-5.71#
27 L6 AR-1254-2	322.431	0.000 E3	100.0#	0#	-5.86#
28 L6 AR-1254-3	570.046	0.000 E3	100.0#	0#	-6.26#
29 L6 AR-1254-4	357.532	0.000 E3	100.0#	0#	-6.49#
30 L6 AR-1254-5	498.269	0.000 E3	100.0#	0#	-6.91#
36 L8 AR-1262-1	256.644	0.000 E3	100.0#	0#	-6.91#
37 L8 AR-1262-2	984.756	0.000 E3	100.0#	0#	-7.45#
38 L8 AR-1262-3	360.087	0.000 E3	100.0#	0#	-7.73#
39 L8 AR-1262-4	715.947	0.000 E3	100.0#	0#	-7.79#
40 L8 AR-1262-5	321.397	0.000 E3	100.0#	0#	-8.30#
41 L9 AR-1268-1	1.122	0.000 E6	100.0#	0#	-7.73#
42 L9 AR-1268-2	1.058	0.000 E6	100.0#	0#	-7.80#
43 L9 AR-1268-3	876.257	0.000 E3	100.0#	0#	-8.01#
44 L9 AR-1268-4	364.105	0.000 E3	100.0#	0#	-8.30#
45 L9 AR-1268-5	3.044	0.000 E6	100.0#	0#	-8.60#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0