

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040323\
 Data File : PQ060791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 01:06
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 06:09:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 27 17:20:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.406	2.763	425.1E6	186.9E6	54.147	47.540
2) SA Decachlor...	8.582	7.540	288.8E6	203.2E6	51.173	43.123
Target Compounds						
26) L6 AR-1254-1	5.327	4.524	170.7E6	108.8E6	536.262	468.941
27) L6 AR-1254-2	5.545	4.671	253.5E6	97030003	514.589	462.248
28) L6 AR-1254-3	5.898	5.052	263.0E6	148.9E6	520.294	450.958
29) L6 AR-1254-4	6.182	5.279	203.4E6	96782673	543.284	462.366
30) L6 AR-1254-5	6.595	5.685	217.4E6	147.1E6	514.120	449.897



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040323\
Data File : PQ060791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 01:06
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 06:09:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 27 17:20:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 425058018
Conc: 54.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.763 min
Delta R.T.: 0.000 min
Response: 186859545
Conc: 47.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.582 min
Delta R.T.: -0.002 min
Response: 288776880
Conc: 51.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.540 min
Delta R.T.: -0.004 min
Response: 203238656
Conc: 43.12 ng/ml



#26 AR-1254-1


R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 170690195
Conc: 536.26 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 108843667
Conc: 468.94 ng/ml




#27 AR-1254-2

R.T.: 5.545 min
Delta R.T.: -0.001 min
Response: 253461108
Conc: 514.59 ng/ml



#27 AR-1254-2


R.T.: 4.671 min
Delta R.T.: -0.003 min
Response: 97030003
Conc: 462.25 ng/ml



#28 AR-1254-3


R.T.: 5.898 min
Delta R.T.: -0.002 min
Response: 263029436
Conc: 520.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.052 min
Delta R.T.: -0.003 min
Response: 148870572
Conc: 450.96 ng/ml




#29 AR-1254-4

R.T.: 6.182 min
Delta R.T.: -0.001 min
Response: 203434819
Conc: 543.28 ng/ml



#29 AR-1254-4


R.T.: 5.279 min
Delta R.T.: -0.003 min
Response: 96782673
Conc: 462.37 ng/ml



#30 AR-1254-5

R.T.: 6.595 min
Delta R.T.: -0.002 min
Response: 217402020
Conc: 514.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.685 min
Delta R.T.: -0.003 min
Response: 147141983
Conc: 449.90 ng/ml