

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040422\
 Data File : PQ056842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2022 19:03
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543074

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/05/2022
 Supervised By :mohammad ahmed 04/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 03:03:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.531	3.749	424.8E6	277.5E6	23.852	25.864
2)	SA Decachlor...	10.301	8.684	731.6E6	457.1E6	50.107	44.311
Target Compounds							
26)	L6 AR-1254-1	6.554	5.601	221.7E6	291.3E6	344.393	436.878 #
27)	L6 AR-1254-2	6.772	5.746	366.3E6	239.1E6	367.200m	435.462
28)	L6 AR-1254-3	7.141	6.144	424.7E6	392.4E6	383.458m	422.207
29)	L6 AR-1254-4	7.429	6.371	272.0E6	290.5E6	371.231m	429.598
30)	L6 AR-1254-5	7.845	6.781	317.8E6	355.8E6	375.712	435.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040422\
Data File : PQ056842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2022 19:03
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543074

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/05/2022
Supervised By :mohammad ahmed 04/05/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 03:03:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

