

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040519\
 Data File : PQ038817.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2019 08:04
 Operator : AJ\SJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248310

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 00:32:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 05 05:12:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.723	4.816	93252157	47836113	19.463	20.998
2)	SA Decachlor...	12.134	10.584	240.2E6	135.1E6	36.311	39.686
Target Compounds							
21)	L5 AR-1248-1	7.169	6.270	42755654	24068280	405.614	399.135
22)	L5 AR-1248-2	7.483	6.560	60749738	34768297	405.752	410.890
23)	L5 AR-1248-3	7.712	6.610	72805676	36374883	398.323	409.580
24)	L5 AR-1248-4	8.158	6.813	78491375	44906336	386.267	391.673
25)	L5 AR-1248-5	8.200	7.268	77385418	44193404	381.324	383.853

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040519\
Data File : PQ038817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2019 08:04
Operator : AJ\SJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1248310

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 00:32:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 05 05:12:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

