

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040621\
Data File : PQ052872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2021 16:08
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_Q
LabSampleId :
AR1254398

Manual Integrations
APPROVED

Ankita
4/8/2021 3:00:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 06:53:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031721CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 18 06:24:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
7.46 581112932 427.65
7.69 886793646 415.49
8.07 941032819 401.47
8.37 663663318 384.21
8.79 818019386 445.88

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.37 365862731 407.96
6.53 321800138 406.23
6.95 520412404 401.26
7.19 346268587 389.12
7.62 477750776 399.37

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040621\
Data File : PQ052872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2021 16:08
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_Q
LabSampleId :
AR1254398

Manual Integrations
APPROVED

Ankita
4/8/2021 3:00:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 06:53:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031721CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 18 06:24:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.46 581112932 427.65
7.69 886793646 415.49
8.07 941032819 401.47
8.37 663663318 384.21
8.79 711831560 388.00
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.37 365862731 407.96
6.53 321800138 406.23
6.95 520412404 401.26
7.19 346268587 389.12
7.62 477750776 399.37

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040621\
 Data File : PQ052872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2021 16:08
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 06:53:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 18 06:24:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.227	4.245	527.5E6	202.7E6	22.890	21.196
2) SA Decachlor...	11.416	9.583	941.4E6	404.5E6	38.068	40.878
Target Compounds						
26) L6 AR-1254-1	7.460	6.372	581.1E6	365.9E6	427.652	407.959
27) L6 AR-1254-2	7.687	6.531	886.8E6	321.8E6	415.492	406.234
28) L6 AR-1254-3	8.071	6.953	941.0E6	520.4E6	401.471	401.263
29) L6 AR-1254-4	8.366	7.191	663.7E6	346.3E6	384.207	389.125
30) L6 AR-1254-5	8.794	7.620	711.8E6	477.8E6	387.998m)	399.373

AJ
 04/14/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Manual Integrations
 APPROVED

Ankita
 4/8/2021 3:00:50 PM

Instrument :
 ECD_Q
 LabSampleId :
 AR1254398