

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040622\
 Data File : PQ056871.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2022 14:49
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543084

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/07/2022
 Supervised By :mohammad ahmed 04/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 01:40:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.535	3.746	390.4E6	256.0E6	21.917	23.866
2)	SA Decachlor...	10.308	8.683	676.7E6	425.7E6	46.345	41.267
Target Compounds							
26)	L6 AR-1254-1	6.557	5.600	206.4E6	272.8E6	320.584	409.021 #
27)	L6 AR-1254-2	6.775	5.744	326.8E6	228.1E6	327.629	415.443 #
28)	L6 AR-1254-3	7.146	6.143	373.6E6	379.3E6	337.310	408.121
29)	L6 AR-1254-4	7.433	6.370	250.3E6	270.9E6	341.671m	400.563
30)	L6 AR-1254-5	7.849	6.780	285.6E6	331.7E6	337.635	406.041

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040622\
Data File : PQ056871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2022 14:49
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543084

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/07/2022
Supervised By :mohammad ahmed 04/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 01:40:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

