

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040720\
 Data File : PQ047302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2020 20:56
 Operator : AJ\MA
 Sample : PQ040720ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 06:35:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 07 06:16:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	50.000	51.277	-2.6	114	0.00
2 SA Decachlorobiphenyl	50.000	51.532	-3.1	117	0.00
3 L1 AR-1016-1	500.000	488.411	2.3	115	0.00
4 L1 AR-1016-2	500.000	498.816	0.2	115	0.00
5 L1 AR-1016-3	500.000	489.596	2.1	114	0.00
6 L1 AR-1016-4	500.000	503.511	-0.7	115	0.00
7 L1 AR-1016-5	500.000	507.356	-1.5	115	0.00
31 L7 AR-1260-1	500.000	498.456	0.3	116	0.00
32 L7 AR-1260-2	500.000	512.526	-2.5	117	0.00
33 L7 AR-1260-3	500.000	524.535	-4.9	119	0.00
34 L7 AR-1260-4	500.000	517.513	-3.5	119	0.00
35 L7 AR-1260-5	500.000	520.512	-4.1	119	0.00

Signal #2

1 SA Tetrachloro-m-xylene	50.000	51.426	-2.9	115	0.00
2 SA Decachlorobiphenyl	50.000	51.946	-3.9	118	0.00
3 L1 AR-1016-1	500.000	498.870	0.2	113	0.00
4 L1 AR-1016-2	500.000	505.690	-1.1	116	0.00
5 L1 AR-1016-3	500.000	512.380	-2.5	114	0.00
6 L1 AR-1016-4	500.000	507.591	-1.5	114	0.00
7 L1 AR-1016-5	500.000	508.501	-1.7	115	0.00
31 L7 AR-1260-1	500.000	510.368	-2.1	115	0.00
32 L7 AR-1260-2	500.000	508.167	-1.6	116	0.00
33 L7 AR-1260-3	500.000	512.782	-2.6	116	0.00
34 L7 AR-1260-4	500.000	516.186	-3.2	116	0.00
35 L7 AR-1260-5	500.000	523.326	-4.7	117	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-5.35#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-5.45#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-5.53#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-5.53#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-6.12#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-6.44#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-6.62#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.71#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-6.42#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-6.44#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-6.51#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040720\
 Data File : PQ047302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2020 20:56
 Operator : AJ\MA
 Sample : PQ040720ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 06:35:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 07 06:16:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-6.62#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-7.40#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-6.42#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-6.71#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-6.93#
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-7.36#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-7.40#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-7.33#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-7.56#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-7.94#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-8.23#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-8.66#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-8.74#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-9.30#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-9.63#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-9.72#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-10.40#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-9.62#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-9.72#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-9.96#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-10.40#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-10.83#

Signal #2

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-4.52#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-4.62#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-4.72#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-4.71#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-5.56#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-5.75#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-5.84#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.03#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-5.53#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-5.56#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-5.75#
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-5.84#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-6.41#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-5.53#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-5.80#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-5.84#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040720\
 Data File : PQ047302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2020 20:56
 Operator : AJ\MA
 Sample : PQ040720ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 06:35:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 07 06:16:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-6.03#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-6.46#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-6.41#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-6.57#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-7.00#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-7.24#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-7.67#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-7.67#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-8.22#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-8.51#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-8.57#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-9.08#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-8.51#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-8.57#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-8.78#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-9.08#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-9.38#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0