

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040722\
 Data File : PQ056906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2022 12:07
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543096

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 15:30:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.531	3.748	392.8E6	257.6E6	22.055	24.008
2) SA Decachlor...	10.303	8.681	702.1E6	440.8E6	48.083	42.727
Target Compounds						
26) L6 AR-1254-1	6.553	5.598	250.3E6	285.2E6	388.770	427.618
27) L6 AR-1254-2	6.770	5.743	362.5E6	239.7E6	363.448	436.568
28) L6 AR-1254-3	7.140	6.141	402.7E6	388.3E6	363.591	417.814
29) L6 AR-1254-4	7.429	6.368	255.6E6	277.3E6	348.847	410.047
30) L6 AR-1254-5	7.845	6.778	285.5E6	339.4E6	337.523	415.455

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ040722\
Data File : PQ056906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2022 12:07
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543096

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 15:30:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

