Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041019\
Data File : PQ039030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2019 22:08
Operator : AJ\SJ
Sample : AR1262CCC400
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262322

Manual Integrations APPROVED
mohammad
4/11/2019 3:58:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 04:55:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 05 05:12:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.709	4.803	104.1E6	54542331	21.736	23.942
2)	SA Decachlor...	12.112	10.566	207.8E6	114.6E6	31.414	33.666
	Target Compounds						
36)	L8 AR-1262-1	9.562	8.556	143.6E6	104.3E6	386.675	398.292
37)	L8 AR-1262-2	10.154	9.079	259.4E6	190.9E6	392.456	399.553
38)	L8 AR-1262-3	10.489	9.372	175.0E6	72456329	387.355m	373.184m
39)	L8 AR-1262-4	10.587	9.439	133.4E6	136.4E6	371.914m	370.396m
40)	L8 AR-1262-5	11.300	9.966	91410456	57078884	367.383	377.868

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041019\
Data File : PQ039030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2019 22:08
Operator : AJ\SJ
Sample : AR1262CCC400
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262322

Manual Integrations APPROVED
mohammad
4/11/2019 3:58:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 04:55:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ040419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 05 05:12:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm