

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041124\
 Data File : PQ066067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2024 21:48
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 22:53:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 12:55:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.448	2.735	220.9E6	337.3E6	44.437	45.484
2) SA Decachlor...	8.659	7.505	121.7E6	281.4E6	41.932	33.670
Target Compounds						
16) L4 AR-1242-1	4.558	3.733	54874033	93125992	511.988	470.283
17) L4 AR-1242-2	4.578	3.748	79736908	133.5E6	513.276	477.208
18) L4 AR-1242-3	4.635	3.908	49902975	73827450	503.137	474.758
19) L4 AR-1242-4	4.728	3.994	39197340	70458656	479.373	471.875
20) L4 AR-1242-5	5.448	4.488	37032975	97245906	454.714	478.626

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041124\
Data File : PQ066067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2024 21:48
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 22:53:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 12:55:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

