

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041125\
 Data File : PQ070211.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2025 18:01
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ041125AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 18:39:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 15:29:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.404	2.783	354.4E6	351.3E6	49.059	50.972
2)	SA Decachlor...	8.501	7.544	429.6E6	662.4E6	88.546	86.310
Target Compounds							
41)	L9 AR-1268-1	7.422	6.510	435.9E6	610.6E6	524.323	525.576
42)	L9 AR-1268-2	7.498	6.572	405.0E6	561.8E6	526.932	527.716
43)	L9 AR-1268-3	7.680	6.773	335.7E6	489.0E6	526.075	526.839
44)	L9 AR-1268-4	7.991	7.060	140.2E6	210.3E6	523.464	528.558
45)	L9 AR-1268-5	8.274	7.327	968.8E6	1452.0E6	528.691	529.493

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041125\
Data File : PQ070211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2025 18:01
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ041125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 18:39:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 15:29:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

