

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ041219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Apr 13 00:36:36 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039044.D 1000 =PQ039042.D 750 =PQ039043.D
 250 =PQ039045.D 50 =PQ039046.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.425	4.610	4.600	4.065	4.459	4.432 E6	4.99	
2) SA	Decachlorobiphenyl	6.822	6.650	6.623	6.898	7.701	6.939 E6	6.36	
3) L1	AR-1016-1	1.784	1.822	1.791	1.706	1.983	1.817 E5	5.60	
4) L1	AR-1016-2	2.683	2.740	2.681	2.481	2.962	2.710 E5	6.35	
5) L1	AR-1016-3	1.670	1.679	1.654	1.554	1.775	1.666 E5	4.73	
6) L1	AR-1016-4	1.377	1.400	1.388	1.274	1.489	1.386 E5	5.53	
7) L1	AR-1016-5	1.409	1.407	1.375	1.353	1.552	1.419 E5	5.48	
8) L2	AR-1221-1	4.367					4.367 E4	0.00	
9) L2	AR-1221-2	3.395					3.395 E4	0.00	
10) L2	AR-1221-3	1.098					1.098 E5	0.00	
11) L3	AR-1232-1	7.954					7.954 E4	0.00	
12) L3	AR-1232-2	4.184					4.184 E4	0.00	
13) L3	AR-1232-3	8.903					8.903 E4	0.00	
14) L3	AR-1232-4	5.124					5.124 E4	0.00	
15) L3	AR-1232-5	4.116					4.116 E4	0.00	
16) L4	AR-1242-1	1.121					1.121 E5	0.00	
17) L4	AR-1242-2	1.642					1.642 E5	0.00	
18) L4	AR-1242-3	1.089					1.089 E5	0.00	
19) L4	AR-1242-4	8.655					8.655 E4	0.00	
20) L4	AR-1242-5	9.879					9.879 E4	0.00	
21) L5	AR-1248-1	8.268					8.268 E4	0.00	
22) L5	AR-1248-2	1.143					1.143 E5	0.00	
23) L5	AR-1248-3	1.371					1.371 E5	0.00	
24) L5	AR-1248-4	1.561					1.561 E5	0.00	
25) L5	AR-1248-5	1.530					1.530 E5	0.00	
26) L6	AR-1254-1	1.722					1.722 E5	0.00	
27) L6	AR-1254-2	2.738					2.738 E5	0.00	
28) L6	AR-1254-3	3.071					3.071 E5	0.00	
29) L6	AR-1254-4	2.221					2.221 E5	0.00	
30) L6	AR-1254-5	2.596					2.596 E5	0.00	
31) L7	AR-1260-1	2.702	2.733	2.676	2.715	3.151	2.795 E5	7.16	
32) L7	AR-1260-2	3.286	3.308	3.251	3.187	3.692	3.345 E5	5.97	
33) L7	AR-1260-3	2.730	2.692	2.641	2.578	3.131	2.754 E5	7.91	
34) L7	AR-1260-4	3.287	3.253	3.157	3.078	3.564	3.268 E5	5.66	
35) L7	AR-1260-5	6.712	6.786	6.591	6.252	6.913	6.651 E5	3.78	
36) L8	AR-1262-1	1.407					1.407 E5	0.00	
37) L8	AR-1262-2	6.341					6.341 E5	0.00	
38) L8	AR-1262-3	4.425					4.425 E5	0.00	
39) L8	AR-1262-4	3.540					3.540 E5	0.00	
40) L8	AR-1262-5	2.528					2.528 E5	0.00	
41) L9	AR-1268-1	8.082					8.082 E5	0.00	
42) L9	AR-1268-2	7.973					7.973 E5	0.00	
43) L9	AR-1268-3	6.594					6.594 E5	0.00	
44) L9	AR-1268-4	2.799					2.799 E5	0.00	
45) L9	AR-1268-5	2.277					2.277 E6	0.00	

Signal #2 Calibration Files

500 =PQ039044.D 1000 =PQ039042.D 750 =PQ039043.D
 250 =PQ039045.D 50 =PQ039046.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.253	2.407	2.363	2.022	2.104	2.230 E6	7.40	
2) SA	Decachlorobiphenyl	3.460	3.391	3.420	3.425	3.750	3.489 E6	4.24	
3) L1	AR-1016-1	1.004	1.036	0.979	0.976	1.083	1.015 E5	4.40	
4) L1	AR-1016-2	1.454	1.533	1.468	1.385	1.566	1.481 E5	4.78	
5) L1	AR-1016-3	7.191	7.769	7.293	7.194	7.859	7.461 E4	4.37	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ041219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Apr 13 00:36:36 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039044.D 1000 =PQ039042.D 750 =PQ039043.D
 250 =PQ039045.D 50 =PQ039046.D

	Compound	500	1000	750	250	50	Avg		%RSD

6)	L1 AR-1016-4	5.823	6.117	5.846	6.133	7.254	6.235	E4	9.43
7)	L1 AR-1016-5	0.835	0.868	0.828	0.898	1.007	0.887	E5	8.19
8)	L2 AR-1221-1	2.376					2.376	E4	0.00
9)	L2 AR-1221-2	2.085					2.085	E4	0.00
10)	L2 AR-1221-3	6.136					6.136	E4	0.00
11)	L3 AR-1232-1	3.792					3.792	E4	0.00
12)	L3 AR-1232-2	4.648					4.648	E4	0.00
13)	L3 AR-1232-3	2.121					2.121	E4	0.00
14)	L3 AR-1232-4	2.105					2.105	E4	0.00
15)	L3 AR-1232-5	2.437					2.437	E4	0.00
16)	L4 AR-1242-1	5.870					5.870	E4	0.00
17)	L4 AR-1242-2	8.594					8.594	E4	0.00
18)	L4 AR-1242-3	4.222					4.222	E4	0.00
19)	L4 AR-1242-4	4.479					4.479	E4	0.00
20)	L4 AR-1242-5	6.485					6.485	E4	0.00
21)	L5 AR-1248-1	4.456					4.456	E4	0.00
22)	L5 AR-1248-2	5.971					5.971	E4	0.00
23)	L5 AR-1248-3	6.401					6.401	E4	0.00
24)	L5 AR-1248-4	8.228					8.228	E4	0.00
25)	L5 AR-1248-5	8.331					8.331	E4	0.00
26)	L6 AR-1254-1	1.429					1.429	E5	0.00
27)	L6 AR-1254-2	1.270					1.270	E5	0.00
28)	L6 AR-1254-3	2.237					2.237	E5	0.00
29)	L6 AR-1254-4	1.403					1.403	E5	0.00
30)	L6 AR-1254-5	2.063					2.063	E5	0.00
31)	L7 AR-1260-1	1.748	1.754	1.704	1.739	2.270	1.843	E5	13.00
32)	L7 AR-1260-2	2.150	2.186	2.131	2.038	2.524	2.206	E5	8.45
33)	L7 AR-1260-3	2.030	2.040	2.016	1.907	2.196	2.038	E5	5.08
34)	L7 AR-1260-4	1.801	1.778	1.787	1.702	2.059	1.825	E5	7.46
35)	L7 AR-1260-5	4.509	4.563	4.583	4.219	4.467	4.468	E5	3.28
36)	L8 AR-1262-1	2.226					2.226	E5	0.00
37)	L8 AR-1262-2	4.318					4.318	E5	0.00
38)	L8 AR-1262-3	1.685					1.685	E5	0.00
39)	L8 AR-1262-4	3.355					3.355	E5	0.00
40)	L8 AR-1262-5	1.438					1.438	E5	0.00
41)	L9 AR-1268-1	5.167					5.167	E5	0.00
42)	L9 AR-1268-2	4.933					4.933	E5	0.00
43)	L9 AR-1268-3	3.985					3.985	E5	0.00
44)	L9 AR-1268-4	1.646					1.646	E5	0.00
45)	L9 AR-1268-5	1.276					1.276	E6	0.00

 (#) = Out of Range