

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041223\
 Data File : PQ060972.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2023 23:17
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603115

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/13/2023
 Supervised By :Ankita Jodhani 04/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 04:19:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 13 01:33:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.405	2.763	165.4E6	73367486	18.872	18.885
2) SA Decachlor...	8.585	7.540	227.2E6	169.3E6	37.899	38.485
Target Compounds						
3) L1 AR-1016-1	4.506	3.766	110.8E6	51934790	375.896	369.241m
4) L1 AR-1016-2	4.525	3.782	163.5E6	78325072	377.315	383.122m
5) L1 AR-1016-3	4.583	3.942	102.7E6	41921891	375.937	388.501
6) L1 AR-1016-4	4.674	3.992	85423594	35300919	378.112	391.647
7) L1 AR-1016-5	4.962	4.186	84961105	43442774	385.956	384.718
31) L7 AR-1260-1	6.064	5.181	157.8E6	88301425	379.988	387.254
32) L7 AR-1260-2	6.324	5.371	188.5E6	108.1E6	376.862	385.341
33) L7 AR-1260-3	6.677	5.513	141.1E6	101.4E6	380.506	381.824
34) L7 AR-1260-4	6.895	5.976	159.5E6	86501081	382.550	386.582
35) L7 AR-1260-5	7.206	6.222	301.9E6	200.2E6	375.929	396.603

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041223\
Data File : PQ060972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2023 23:17
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603115

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/13/2023
Supervised By :Ankita Jodhani 04/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 04:19:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 13 01:33:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

