

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041422\
Data File : PQ056974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2022 17:34
Operator : YP\AJ
Sample : N2290-01
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
BFFC8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/15/2022
Supervised By :mohammad ahmed 04/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:58:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.527min 16.743 ng/ml

response 298220060

(1) Tetrachloro-m-xylene #2 (SA)

3.744min 17.465 ng/ml

response 187365597

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041422\
Data File : PQ056974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2022 17:34
Operator : YP\AJ
Sample : N2290-01
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
BFFC8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/15/2022
Supervised By :mohammad ahmed 04/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:58:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.526min 16.258 ng/ml m

response 289571791

(1) Tetrachloro-m-xylene #2 (SA)

3.744min 17.465 ng/ml

response 187365597

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041422\
 Data File : PQ056974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2022 17:34
 Operator : YP\AJ
 Sample : N2290-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1



Instrument :
 ECD_Q
 ClientSampleId :
 BFFC8

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:58:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 04/15/2022
 Supervised By :mohammad ahmed 04/15/2022

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 7.29 25479452 39.76
 7.55 47167940 64.00
 7.84 65379947 70.10
 8.14 45261711 65.56
 8.47 61854342 44.56

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 6.26 23491212 43.00
 6.45 27769092 41.73
 6.60 17830936 28.82
 7.07 16440316 31.09
 7.31 56913886 43.48

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041422\
 Data File : PQ056974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2022 17:34
 Operator : YP\AJ
 Sample : N2290-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1



Instrument :
 ECD_Q
 ClientSampleId :
 BFFC8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/15/2022
 Supervised By :mohammad ahmed 04/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:58:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 7.29 23397850 36.51
 7.55 47167940 64.00
 7.84 56107887 60.16
 8.14 31029326 44.94
 8.47 53403102 38.47

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 6.26 23491212 43.00
 6.45 27769092 41.73
 6.60 23645543 38.21
 7.07 16440316 31.09
 7.31 56913886 43.48