

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041522\
 Data File : PQ056987.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Apr 2022 15:03
 Operator : YP\AJ
 Sample : N2280-09
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 BFFC7

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/18/2022
 Supervised By :mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 23:31:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 24 09:20:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.528	3.745	313.3E6	209.9E6	17.590m	19.568
2) SA Decachlor...	10.298	8.677	423.4E6	245.4E6	28.999	23.785
Target Compounds						
26) L6 AR-1254-1	6.556	5.597	51768986	27088474	80.411m	40.622m#
27) L6 AR-1254-2	6.770	5.741	57448807	20782050	57.593m	37.850 #
28) L6 AR-1254-3	7.137	6.139	41316908	42733062	37.306m	45.979m
29) L6 AR-1254-4	7.425	6.364	27868822	15258337	38.042m	22.563m#
30) L6 AR-1254-5	7.838	6.774	28267155	55290787	33.418m	67.682m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041522\
Data File : PQ056987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Apr 2022 15:03
Operator : YP\AJ
Sample : N2280-09
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
BFFC7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2022
Supervised By :mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 23:31:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032322CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 24 09:20:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

