

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041624\
 Data File : PQ066107.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2024 09:55
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 13:10:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 12:55:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.443	2.735	207.1E6	317.5E6	41.669	42.814
2)	SA Decachlor...	8.651	7.500	123.5E6	311.2E6	42.573	37.241
Target Compounds							
26)	L6 AR-1254-1	5.379	4.486	70538443	170.7E6	465.395	424.818
27)	L6 AR-1254-2	5.597	4.633	107.6E6	155.8E6	467.467	423.955
28)	L6 AR-1254-3	5.952	5.014	109.5E6	239.3E6	462.841	412.754
29)	L6 AR-1254-4	6.237	5.240	80369538	157.0E6	459.769	403.290
30)	L6 AR-1254-5	6.651	5.646	88252008	231.7E6	457.411	405.892



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041624\
Data File : PQ066107.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2024 09:55
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 13:10:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 12:55:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.443 min
Delta R.T.: 0.000 min
Response: 207105525
Conc: 41.67 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.735 min
Delta R.T.: -0.001 min
Response: 317483473
Conc: 42.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.651 min
Delta R.T.: -0.001 min
Response: 123540171
Conc: 42.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.500 min
Delta R.T.: -0.005 min
Response: 311222879
Conc: 37.24 ng/ml



#26 AR-1254-1


R.T.: 5.379 min
Delta R.T.: 0.001 min
Response: 70538443
Conc: 465.40 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.486 min
Delta R.T.: -0.003 min
Response: 170748555
Conc: 424.82 ng/ml




#27 AR-1254-2

R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 107596496
Conc: 467.47 ng/ml



#27 AR-1254-2


R.T.: 4.633 min
Delta R.T.: -0.003 min
Response: 155846368
Conc: 423.95 ng/ml



#28 AR-1254-3


R.T.: 5.952 min
Delta R.T.: 0.000 min
Response: 109541264
Conc: 462.84 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.014 min
Delta R.T.: -0.004 min
Response: 239327740
Conc: 412.75 ng/ml




#29 AR-1254-4

R.T.: 6.237 min
Delta R.T.: 0.001 min
Response: 80369538
Conc: 459.77 ng/ml



#29 AR-1254-4


R.T.: 5.240 min
Delta R.T.: -0.004 min
Response: 157022811
Conc: 403.29 ng/ml



#30 AR-1254-5

R.T.: 6.651 min
Delta R.T.: 0.000 min
Response: 88252008
Conc: 457.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.646 min
Delta R.T.: -0.004 min
Response: 231675064
Conc: 405.89 ng/ml