

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042019\
 Data File : PQ039189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2019 13:12
 Operator : AJ\SJ
 Sample : K2241-08
 Misc : AR1248 LOQ 100 PPB
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT2-2019

Manual Integrations
 APPROVED

mohammad
 4/29/2019 4:11:11 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 00:35:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 20 00:43:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.706	4.795	96625605	57143296	23.356	22.730
2)	SA Decachlor...	12.116	10.563	109.0E6	63261257	18.427	18.353
Target Compounds							
21)	L5 AR-1248-1	7.155	6.249	9017627	5594609	95.301	89.022m
22)	L5 AR-1248-2	7.471	6.540	13091340	8213852	100.677m	96.186
23)	L5 AR-1248-3	7.699	6.589	15324991	8789697	97.888	97.518
24)	L5 AR-1248-4	8.147	6.793	16040237	10296890	89.542	92.643
25)	L5 AR-1248-5	8.188	7.249	15807274	8654883	89.056	76.451

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042019\
Data File : PQ039189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2019 13:12
Operator : AJ\SJ
Sample : K2241-08
Misc : AR1248 LOQ 100 PPB
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT2-2019

Manual Integrations
APPROVED

mohammad
4/29/2019 4:11:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 00:35:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042019.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Apr 20 00:43:12 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

