

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ042019.M
 Title : GC EXTRACTABLES
 Last Update : Sat Apr 20 00:43:12 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039129.D 1000 =PQ039127.D 750 =PQ039128.D
 250 =PQ039130.D 50 =PQ039131.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.797	4.950	4.197	3.765	3.976	4.137 E6	11.74
2) SA	Decachlorobiphenyl	5.735	5.491	5.668	5.899	6.778	5.914 E6	8.53
3) L1	AR-1016-1	1.384	1.684	1.494	1.440	1.746	1.550 E5	10.17
4) L1	AR-1016-2	2.090	2.504	2.255	2.164	2.597	2.322 E5	9.44
5) L1	AR-1016-3	1.265	1.480	1.355	1.320	1.492	1.382 E5	7.23
6) L1	AR-1016-4	1.078	1.238	1.124	1.082	1.352	1.175 E5	10.06
7) L1	AR-1016-5	1.096	1.232	1.154	1.111	1.337	1.186 E5	8.42
8) L2	AR-1221-1	4.781					4.781 E4	0.00
9) L2	AR-1221-2	3.461					3.461 E4	0.00
10) L2	AR-1221-3	1.144					1.144 E5	0.00
11) L3	AR-1232-1	8.791					8.791 E4	0.00
12) L3	AR-1232-2	4.573					4.573 E4	0.00
13) L3	AR-1232-3	9.228					9.228 E4	0.00
14) L3	AR-1232-4	4.600					4.600 E4	0.00
15) L3	AR-1232-5	3.363					3.363 E4	0.00
16) L4	AR-1242-1	1.380					1.380 E5	0.00
17) L4	AR-1242-2	2.002					2.002 E5	0.00
18) L4	AR-1242-3	1.226					1.226 E5	0.00
19) L4	AR-1242-4	1.018					1.018 E5	0.00
20) L4	AR-1242-5	1.148					1.148 E5	0.00
21) L5	AR-1248-1	9.462					9.462 E4	0.00
22) L5	AR-1248-2	1.300					1.300 E5	0.00
23) L5	AR-1248-3	1.566					1.566 E5	0.00
24) L5	AR-1248-4	1.791					1.791 E5	0.00
25) L5	AR-1248-5	1.775					1.775 E5	0.00
26) L6	AR-1254-1	2.065					2.065 E5	0.00
27) L6	AR-1254-2	3.261					3.261 E5	0.00
28) L6	AR-1254-3	3.552					3.552 E5	0.00
29) L6	AR-1254-4	2.485					2.485 E5	0.00
30) L6	AR-1254-5	2.811					2.811 E5	0.00
31) L7	AR-1260-1	2.156	2.275	2.192	2.230	2.471	2.265 E5	5.46
32) L7	AR-1260-2	2.597	2.720	2.637	2.650	3.093	2.739 E5	7.40
33) L7	AR-1260-3	2.114	2.164	2.100	2.151	2.628	2.231 E5	10.02
34) L7	AR-1260-4	2.464	2.518	2.449	2.555	3.306	2.658 E5	13.71
35) L7	AR-1260-5	4.953	5.181	5.085	4.950	5.465	5.127 E5	4.14
36) L8	AR-1262-1	1.391					1.391 E5	0.00
37) L8	AR-1262-2	6.298					6.298 E5	0.00
38) L8	AR-1262-3	4.277					4.277 E5	0.00
39) L8	AR-1262-4	3.401					3.401 E5	0.00
40) L8	AR-1262-5	2.302					2.302 E5	0.00
41) L9	AR-1268-1	7.543					7.543 E5	0.00
42) L9	AR-1268-2	7.184					7.184 E5	0.00
43) L9	AR-1268-3	6.142					6.142 E5	0.00
44) L9	AR-1268-4	2.479					2.479 E5	0.00
45) L9	AR-1268-5	2.089					2.089 E6	0.00

Signal #2 Calibration Files

500 =PQ039129.D 1000 =PQ039127.D 750 =PQ039128.D
 250 =PQ039130.D 50 =PQ039131.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.285	3.080	2.518	2.210	2.478	2.514 E6	13.58
2) SA	Decachlorobiphenyl	3.347	3.317	3.408	3.383	3.781	3.447 E6	5.50
3) L1	AR-1016-1	0.923	1.137	0.977	0.957	0.996	0.998 E5	8.23
4) L1	AR-1016-2	1.324	1.649	1.412	1.324	1.451	1.432 E5	9.33
5) L1	AR-1016-3	7.222	8.567	7.465	7.137	8.576	7.793 E4	9.24

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ042019.M
 Title : GC EXTRACTABLES
 Last Update : Sat Apr 20 00:43:12 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039129.D 1000 =PQ039127.D 750 =PQ039128.D
 250 =PQ039130.D 50 =PQ039131.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	5.792	6.668	5.881	5.771	6.947	6.212	E4	8.92
7)	L1 AR-1016-5	7.966	9.049	8.112	7.874	8.960	8.392	E4	6.75
8)	L2 AR-1221-1	2.749					2.749	E4	0.00
9)	L2 AR-1221-2	2.032					2.032	E4	0.00
10)	L2 AR-1221-3	6.875					6.875	E4	0.00
11)	L3 AR-1232-1	5.109					5.109	E4	0.00
12)	L3 AR-1232-2	5.886					5.886	E4	0.00
13)	L3 AR-1232-3	3.123					3.123	E4	0.00
14)	L3 AR-1232-4	2.839					2.839	E4	0.00
15)	L3 AR-1232-5	3.395					3.395	E4	0.00
16)	L4 AR-1242-1	8.900					8.900	E4	0.00
17)	L4 AR-1242-2	1.261					1.261	E5	0.00
18)	L4 AR-1242-3	6.873					6.873	E4	0.00
19)	L4 AR-1242-4	6.859					6.859	E4	0.00
20)	L4 AR-1242-5	9.047					9.047	E4	0.00
21)	L5 AR-1248-1	6.285					6.285	E4	0.00
22)	L5 AR-1248-2	8.540					8.540	E4	0.00
23)	L5 AR-1248-3	9.013					9.013	E4	0.00
24)	L5 AR-1248-4	1.111					1.111	E5	0.00
25)	L5 AR-1248-5	1.132					1.132	E5	0.00
26)	L6 AR-1254-1	2.049					2.049	E5	0.00
27)	L6 AR-1254-2	1.772					1.772	E5	0.00
28)	L6 AR-1254-3	3.061					3.061	E5	0.00
29)	L6 AR-1254-4	1.835					1.835	E5	0.00
30)	L6 AR-1254-5	2.616					2.616	E5	0.00
31)	L7 AR-1260-1	1.608	1.741	1.635	1.629	1.877	1.698	E5	6.63
32)	L7 AR-1260-2	1.951	2.125	1.997	1.971	2.206	2.050	E5	5.38
33)	L7 AR-1260-3	1.872	2.026	1.906	1.869	2.078	1.950	E5	4.92
34)	L7 AR-1260-4	1.604	1.700	1.628	1.603	1.756	1.658	E5	4.06
35)	L7 AR-1260-5	3.998	4.275	4.129	3.932	4.095	4.086	E5	3.22
36)	L8 AR-1262-1	2.707					2.707	E5	0.00
37)	L8 AR-1262-2	5.079					5.079	E5	0.00
38)	L8 AR-1262-3	1.900					1.900	E5	0.00
39)	L8 AR-1262-4	3.711					3.711	E5	0.00
40)	L8 AR-1262-5	1.568					1.568	E5	0.00
41)	L9 AR-1268-1	5.751					5.751	E5	0.00
42)	L9 AR-1268-2	5.384					5.384	E5	0.00
43)	L9 AR-1268-3	4.374					4.374	E5	0.00
44)	L9 AR-1268-4	1.761					1.761	E5	0.00
45)	L9 AR-1268-5	1.384					1.384	E6	0.00

(#) = Out of Range