

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042222\
 Data File : PQ057124.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2022 01:34
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ042222AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 10:13:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Apr 24 10:06:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.525	3.741	1049.9E6	630.9E6	51.113	54.928
2) SA Decachlor...	10.288	8.668	1091.1E6	571.8E6	53.671	55.995
Target Compounds						
26) L6 AR-1254-1	6.547	5.592	292.3E6	326.6E6	482.493	490.265
27) L6 AR-1254-2	6.765	5.737	454.8E6	271.4E6	473.816	485.709
28) L6 AR-1254-3	7.135	6.135	531.1E6	464.5E6	481.763	493.942
29) L6 AR-1254-4	7.423	6.362	354.2E6	346.3E6	462.149	492.527
30) L6 AR-1254-5	7.838	6.771	401.7E6	414.6E6	454.495	484.719

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042222\
Data File : PQ057124.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2022 01:34
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042222AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 10:13:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Apr 24 10:06:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

