

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042324\
 Data File : PQ066219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2024 16:16
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 03:59:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 03:59:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.440	2.734	323.3E6	505.1E6	75.863	75.905
2) SA Decachlor...	8.650	7.500	326.6E6	893.7E6	75.489	75.850
Target Compounds						
41) L9 AR-1268-1	7.534	6.461	310.9E6	827.2E6	754.460	760.336
42) L9 AR-1268-2	7.612	6.524	287.0E6	757.9E6	754.529	757.969
43) L9 AR-1268-3	7.794	6.723	238.3E6	658.0E6	755.175	757.266
44) L9 AR-1268-4	8.117	7.016	103.0E6	277.9E6	755.448	757.165
45) L9 AR-1268-5	8.412	7.284	706.3E6	1988.4E6	757.136	761.406



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042324\
Data File : PQ066219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2024 16:16
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 03:59:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 03:59:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.440 min
Delta R.T.: 0.000 min
Response: 323327224
Conc: 75.86 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.734 min
Delta R.T.: 0.000 min
Response: 505136929
Conc: 75.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: 0.000 min
Response: 326649781
Conc: 75.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.500 min
Delta R.T.: 0.000 min
Response: 893698979
Conc: 75.85 ng/ml



#41 AR-1268-1


R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 310927935
Conc: 754.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 827216167
Conc: 760.34 ng/ml




#42 AR-1268-2

R.T.: 7.612 min
Delta R.T.: 0.000 min
Response: 286981372
Conc: 754.53 ng/ml



#42 AR-1268-2


R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 757892630
Conc: 757.97 ng/ml



#43 AR-1268-3


R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 238274280
Conc: 755.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 658024623
Conc: 757.27 ng/ml




#44 AR-1268-4

R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 102988786
Conc: 755.45 ng/ml



#44 AR-1268-4


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 277887039
Conc: 757.16 ng/ml



#45 AR-1268-5

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 706271305
Conc: 757.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 7.284 min
Delta R.T.: 0.000 min
Response: 1988435065
Conc: 761.41 ng/ml