

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042324\
 Data File : PQ066227.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2024 18:15
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ042324

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 02:22:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 02:17:34 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	2.734	224.9E6	349.4E6	51.300	50.902
2)	SA Decachlor...	8.650	7.500	225.3E6	613.3E6	50.281	50.478
Target Compounds							
41)	L9 AR-1268-1	7.534	6.461	214.4E6	560.4E6	503.081	496.857
42)	L9 AR-1268-2	7.613	6.524	196.3E6	518.0E6	496.986	502.771
43)	L9 AR-1268-3	7.794	6.722	164.0E6	449.8E6	503.570	501.728
44)	L9 AR-1268-4	8.117	7.016	71479967	191.3E6	504.126	505.117
45)	L9 AR-1268-5	8.413	7.284	479.6E6	1347.5E6	502.867	505.825



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042324\
Data File : PQ066227.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2024 18:15
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042324

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 02:22:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 02:17:34 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.441 min
Delta R.T.: 0.000 min
Response: 224915632
Conc: 51.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042324




#1 Tetrachloro-m-xylene

R.T.: 2.734 min
Delta R.T.: 0.000 min
Response: 349364909
Conc: 50.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: 0.000 min
Response: 225303739
Conc: 50.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.500 min
Delta R.T.: 0.000 min
Response: 613311022
Conc: 50.48 ng/ml



#41 AR-1268-1


R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 214435512
Conc: 503.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042324




#41 AR-1268-1

R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 560429837
Conc: 496.86 ng/ml




#42 AR-1268-2

R.T.: 7.613 min
Delta R.T.: 0.000 min
Response: 196332249
Conc: 496.99 ng/ml



#42 AR-1268-2


R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 518045740
Conc: 502.77 ng/ml



#43 AR-1268-3


R.T.: 7.794 min
Delta R.T.: 0.000 min
Response: 163950247
Conc: 503.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042324




#43 AR-1268-3

R.T.: 6.722 min
Delta R.T.: 0.000 min
Response: 449809457
Conc: 501.73 ng/ml




#44 AR-1268-4

R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 71479967
Conc: 504.13 ng/ml



#44 AR-1268-4


R.T.: 7.016 min
Delta R.T.: -0.001 min
Response: 191326430
Conc: 505.12 ng/ml



#45 AR-1268-5

R.T.: 8.413 min
Delta R.T.: 0.000 min
Response: 479607364
Conc: 502.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ042324



#45 AR-1268-5

R.T.: 7.284 min
Delta R.T.: 0.000 min
Response: 1347529929
Conc: 505.83 ng/ml