

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042424\
 Data File : PQ066239.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 12:44
 Operator : YP\AJ
 Sample : PB160460BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ALCS460

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/25/2024
 Supervised By :Ankita Jodhani 04/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 21:19:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 04:31:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	2.733	88276690	131.7E6	19.408	20.059
2)	SA Decachlor...	8.650	7.499	56117309	146.6E6	43.039	45.414
Target Compounds							
3)	L1 AR-1016-1	4.551	3.732	14798042	25384244	104.298	106.012
4)	L1 AR-1016-2	4.570	3.746	22519306	37775780	103.090	108.205
5)	L1 AR-1016-3	4.628	3.907	14923687	20130100	106.179	106.963
6)	L1 AR-1016-4	4.721	3.956	12004684	17508364	104.238	110.670
7)	L1 AR-1016-5	5.009	4.150	11605215	20774983	107.429	105.855
31)	L7 AR-1260-1	6.115	5.142	20626662	42936445	109.044	109.306
32)	L7 AR-1260-2	6.376	5.333	23780899	52049496	107.240	108.504
33)	L7 AR-1260-3	6.730	5.473	14495976	48532276	89.491	108.465
34)	L7 AR-1260-4	6.948	5.935	17866665	35223565	96.947m	95.865
35)	L7 AR-1260-5	7.260	6.182	33201918	79869628	97.362	94.810

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042424\
Data File : PQ066239.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 12:44
Operator : YP\AJ
Sample : PB160460BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ALCS460

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2024
Supervised By :Ankita Jodhani 04/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 21:19:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 04:31:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

