

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042424\
 Data File : PQ066248.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 14:56
 Operator : YP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12483083

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 21:22:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 04:31:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	2.734	96465563	148.4E6	21.208	22.602
2)	SA Decachlor...	8.649	7.500	57532972	152.4E6	44.125	47.197
Target Compounds							
21)	L5 AR-1248-1	4.551	3.732	35329561	61004399	418.094	421.920
22)	L5 AR-1248-2	4.818	3.956	51550308	92819197	429.883	422.994
23)	L5 AR-1248-3	5.009	3.992	62267248	88610961	441.018	419.766
24)	L5 AR-1248-4	5.405	4.150	62404409	110.0E6	419.968	423.255
25)	L5 AR-1248-5	5.440	4.524	58878773	112.2E6	390.718	434.670

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042424\
Data File : PQ066248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 14:56
Operator : YP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12483083

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 21:22:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 04:31:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

