

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042522\
 Data File : PQ057150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2022 22:51
 Operator : YP\AJ
 Sample : AR1248ICC400
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12483013

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/26/2022
 Supervised By :mohammad ahmed 04/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 06:34:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 26 06:33:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.525	3.741	473.7E6	268.4E6	20.000	20.000
2) SA Decachlor...	10.291	8.669	911.9E6	485.0E6	40.000	40.000
Target Compounds						
21) L5 AR-1248-1	5.701	4.807	130.3E6	97786291	400.000	400.000
22) L5 AR-1248-2	5.967	5.036	204.1E6	146.2E6	400.000	400.000
23) L5 AR-1248-3	6.173	5.078	251.3E6	153.9E6	400.000	400.000
24) L5 AR-1248-4	6.579	5.246	272.7E6	184.4E6	400.000	400.000
25) L5 AR-1248-5	6.619	5.634	292.3E6	192.8E6	400.000	397.145m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ042522\
Data File : PQ057150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2022 22:51
Operator : YP\AJ
Sample : AR1248ICC400
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12483013

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/26/2022
Supervised By :mohammad ahmed 04/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 06:34:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 26 06:33:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

