

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050419\
 Data File : PQ039376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 May 2019 11:49
 Operator : AJ\SJ
 Sample : K2613-19
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 BFHD3

Manual Integrations
 APPROVED

Ankita
 5/6/2019 10:13:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 23:45:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 02 08:13:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.695	4.793	35088234	21219547	11.580	12.804
2) SA Decachlor...	12.090	10.547	154.2E6	76077181	25.242m	22.549m
Target Compounds						
31) L7 AR-1260-1	8.908	7.950	122.4E6	83674141	371.854	354.191
32) L7 AR-1260-2	9.178	8.153	159.1E6	110.1E6	402.204	372.276
33) L7 AR-1260-3	9.553	8.318	191.0E6	105.1E6	621.080	376.996 #
34) L7 AR-1260-4	9.796	8.816	233.8E6	131.7E6	644.885	554.955
35) L7 AR-1260-5	10.144	9.067	511.9E6	372.0E6	672.406	621.940

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050419\
Data File : PQ039376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 May 2019 11:49
Operator : AJ\SJ
Sample : K2613-19
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
BFHD3

Manual Integrations
APPROVED

Ankita
5/6/2019 10:13:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 23:45:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 02 08:13:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

