

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050422\
 Data File : PQ057376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2022 14:13
 Operator : YP\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12423204

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 02:30:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 26 08:09:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.525	3.738	526.6E6	262.5E6	25.078	22.055
2)	SA Decachlor...	10.291	8.664	830.6E6	433.8E6	40.411	42.085
Target Compounds							
16)	L4 AR-1242-1	5.702	4.805	174.9E6	113.2E6	433.377	381.827
17)	L4 AR-1242-2	5.725	4.823	376.5E6	217.1E6	464.613	408.631
18)	L4 AR-1242-3	5.785	4.996	253.2E6	94691559	477.977	394.149
19)	L4 AR-1242-4	5.888	5.075	200.9E6	98167249	477.494	410.865
20)	L4 AR-1242-5	6.620	5.590	199.3E6	126.2E6	475.889	399.376

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050422\
Data File : PQ057376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2022 14:13
Operator : YP\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12423204

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 02:30:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 26 08:09:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

