

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050619\
Data File : PQ039432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2019 13:15
Operator : AJ\SJ
Sample : K2613-08DL2 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
BFH50DL2


Manual Integrations
APPROVED

Ankita

5/7/2019 10:25:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 00:32:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 02 08:13:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.698min 0.815 ng/ml

response 2470365

(1) Tetrachloro-m-xylene #2 (SA)

4.793min 0.721 ng/ml

response 1194162

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050619\
Data File : PQ039432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2019 13:15
Operator : AJ\SJ
Sample : K2613-08DL2 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
BFH50DL2


Manual Integrations
APPROVED

Ankita

5/7/2019 10:25:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 00:32:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 02 08:13:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.697min 0.617 ng/ml m

response 1869076

(1) Tetrachloro-m-xylene #2 (SA)

4.793min 0.715 ng/ml m

response 1184437

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050619\
Data File : PQ039432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2019 13:15
Operator : AJ\SJ
Sample : K2613-08DL2 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
BFH50DL2


Manual Integrations
APPROVED

Ankita

5/7/2019 10:25:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 00:32:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 02 08:13:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

12.101min 1.332 ng/ml

response 8136371

(2) Decachlorobiphenyl #2 (SA)

10.552min 1.026 ng/ml

response 3460265

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050619\
Data File : PQ039432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2019 13:15
Operator : AJ\SJ
Sample : K2613-08DL2 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
BFH50DL2


Manual Integrations
APPROVED

Ankita

5/7/2019 10:25:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 00:32:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 02 08:13:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

12.100min 1.344 ng/ml m

response 8209112

(2) Decachlorobiphenyl #2 (SA)

10.553min 1.305 ng/ml m

response 4402198

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050619\
 Data File : PQ039432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2019 13:15
 Operator : AJ\SJ
 Sample : K2613-08DL2 20X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 BFH50DL2

Manual Integrations
 APPROVED

Ankita

5/7/2019 10:25:52 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 00:32:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 02 08:13:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.697	4.793	1869076	1184437	0.617m	0.715m
2) SA Decachlor...	12.100	10.553	8209112	4402198	1.344m	1.305m
Target Compounds						
31) L7 AR-1260-1	8.913	7.952	297.0E6	220.5E6	902.422	933.353
32) L7 AR-1260-2	9.184	8.156	432.8E6	331.5E6	1094.109	1120.863
33) L7 AR-1260-3	9.560	8.321	385.7E6	273.3E6	1254.488	980.705
34) L7 AR-1260-4	9.803	8.819	450.5E6	308.4E6	1242.461	1300.103
35) L7 AR-1260-5	10.151	9.071	1148.4E6	912.6E6	1508.315	1525.766

AJ
 05/08/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.