

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050622\
 Data File : PQ057465.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2022 15:15
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543215

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 05:58:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 06 23:16:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.521	3.738	714.3E6	275.9E6	20.905	20.747
2) SA Decachlor...	10.282	8.659	1235.2E6	465.7E6	41.535	42.818
Target Compounds						
26) L6 AR-1254-1	6.544	5.587	358.4E6	289.2E6	402.449	380.107
27) L6 AR-1254-2	6.761	5.731	567.1E6	245.3E6	404.120	398.730
28) L6 AR-1254-3	7.131	6.129	664.4E6	404.1E6	396.631	391.138
29) L6 AR-1254-4	7.419	6.357	436.9E6	294.7E6	396.791	388.015
30) L6 AR-1254-5	7.835	6.764	487.3E6	355.9E6	391.181	381.480

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050622\
Data File : PQ057465.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2022 15:15
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543215

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 05:58:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 06 23:16:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

