

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050923\
 Data File : PQ061159.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 20:50
 Operator : YP\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 01:24:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050923CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 10 01:22:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.405	2.761	420.8E6	247.2E6	72.005	73.297
2) SA Decachlor...	8.585	7.535	610.2E6	563.0E6	138.385	141.679
Target Compounds						
21) L5 AR-1248-1	4.506	3.765	173.1E6	109.0E6	1363.623	1392.612
22) L5 AR-1248-2	4.771	3.990	229.6E6	156.7E6	1343.628	1363.224
23) L5 AR-1248-3	4.961	4.026	281.9E6	151.5E6	1360.273	1374.112
24) L5 AR-1248-4	5.355	4.184	327.9E6	189.1E6	1384.272	1380.499
25) L5 AR-1248-5	5.390	4.559	323.0E6	195.2E6	1390.977	1414.088



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050923\
Data File : PQ061159.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 20:50
Operator : YP\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:24:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050923CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 10 01:22:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.405 min
Delta R.T.: 0.000 min
Response: 420781726
Conc: 72.01 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.761 min
Delta R.T.: 0.000 min
Response: 247175712
Conc: 73.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.585 min
Delta R.T.: 0.000 min
Response: 610179624
Conc: 138.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 563048270
Conc: 141.68 ng/ml



#21 AR-1248-1


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 173080680
Conc: 1363.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#21 AR-1248-1

R.T.: 3.765 min
Delta R.T.: 0.000 min
Response: 108981109
Conc: 1392.61 ng/ml




#22 AR-1248-2

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 229636650
Conc: 1343.63 ng/ml



#22 AR-1248-2


R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 156667981
Conc: 1363.22 ng/ml



#23 AR-1248-3


R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 281858619
Conc: 1360.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#23 AR-1248-3

R.T.: 4.026 min
Delta R.T.: 0.000 min
Response: 151476175
Conc: 1374.11 ng/ml




#24 AR-1248-4

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 327905320
Conc: 1384.27 ng/ml



#24 AR-1248-4


R.T.: 4.184 min
Delta R.T.: 0.000 min
Response: 189058660
Conc: 1380.50 ng/ml



#25 AR-1248-5

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 323035880
Conc: 1390.98 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#25 AR-1248-5

R.T.: 4.559 min
Delta R.T.: 0.000 min
Response: 195176559
Conc: 1414.09 ng/ml