

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050924\
 Data File : PQ066448.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2024 20:28
 Operator : YP\AJ
 Sample : P2403-07
 Misc : AR1268 LOD 40 PPB
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 01:16:48 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 04:24:21 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.440	2.733	94169338	140.6E6	21.662	21.983
2) SA Decachlor...	8.649	7.497	62312200	167.1E6	25.317	26.562
Target Compounds						
3) L1 AR-1016-1	4.567	3.732	469584	135308	3.542	0.580 #
4) L1 AR-1016-2	4.567	3.747	469584	282377	2.345	0.834 #
5) L1 AR-1016-3	4.626	3.907	242732	147162	1.935	0.785 #
6) L1 AR-1016-4	4.726	3.957	188626	476143	1.887	3.019 #
7) L1 AR-1016-5	0.000	4.149	0	254393	N.D.	1.311 #
8) L2 AR-1221-1	0.000	2.931	0	255424	N.D.	3.032 #
9) L2 AR-1221-2	3.721	3.000	1320996	1961475	34.184	32.428
10) L2 AR-1221-3	3.796	3.076	190743	212290	1.558	1.134 #
11) L3 AR-1232-1	3.796	3.076	190743	212290	1.889	1.379 #
12) L3 AR-1232-2	4.279	3.747	770317	282377	13.973	1.752 #
13) L3 AR-1232-3	4.567	3.907	469584	147162	4.871	1.669 #
14) L3 AR-1232-4	4.726	3.988	188626	705366	3.985	9.362 #
15) L3 AR-1232-5	4.823	4.149	437928	254393	13.153	3.017 #
16) L4 AR-1242-1	4.567	3.732	469584	135308	4.239	0.682 #
17) L4 AR-1242-2	4.567	3.747	469584	282377	2.779	0.998 #
18) L4 AR-1242-3	4.626	3.907	242732	147162	2.293	0.938 #
19) L4 AR-1242-4	4.726	3.988	188626	705366	2.262	4.808 #
20) L4 AR-1242-5	5.450	4.485	265623	654157	3.007	3.402
21) L5 AR-1248-1	4.567	3.732	469584	135308	5.810	0.909 #
22) L5 AR-1248-2	4.823	3.957	437928	476143	3.798	2.152 #
23) L5 AR-1248-3	0.000	3.988	0	705366	N.D.	3.335 #
24) L5 AR-1248-4	5.402	4.149	2763472	254393	19.176	0.988 #
25) L5 AR-1248-5	5.450	4.509	265623	687344	1.817	2.663 #
26) L6 AR-1254-1	5.402f	4.485	2763472	654157	18.394	1.663 #
27) L6 AR-1254-2	5.592	4.628	364198	606215	1.545	1.779
28) L6 AR-1254-3	5.952	5.027	212511	4976660	0.865	9.215 #
29) L6 AR-1254-4	6.248	5.246	1517436	709861	8.713	1.911 #
30) L6 AR-1254-5	6.638	5.646	302838	296006	1.649	0.608 #
31) L7 AR-1260-1	6.109	5.142	500294	647041	2.765	1.652 #
32) L7 AR-1260-2	6.418f	5.327	246966	1472830	1.165	3.090 #
33) L7 AR-1260-3	6.730	5.473	2491259	1069616	16.199	2.430 #
34) L7 AR-1260-4	6.970	5.933	3742460	8201868	21.520	21.542
35) L7 AR-1260-5	7.259	6.181	1211875	3776369	3.701	4.610
36) L8 AR-1262-1	6.638	5.646	302838	296006	2.062	1.060 #
37) L8 AR-1262-2	7.259	5.933	1211875	8201868	3.307	17.250 #
38) L8 AR-1262-3	7.534	6.458	12786872	33280187	53.141	88.116 #
39) L8 AR-1262-4	7.612	6.521	11810023	28727886	63.613	41.162 #
40) L8 AR-1262-5	8.116	7.014	4097951	10072684	33.130	30.606
41) L9 AR-1268-1	7.534	6.458	12786872	33280187	30.524	29.701

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050924\
Data File : PQ066448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2024 20:28
Operator : YP\AJ
Sample : P2403-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:16:48 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 04:24:21 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	7.612	6.521	11810023	28727886	30.363	28.160
43)	L9 AR-1268-3	7.793	6.720	9937744	25338310	30.919	28.485
44)	L9 AR-1268-4	8.116	7.014	4097951	10072684	28.920	27.083
45)	L9 AR-1268-5	8.413	7.282	27342077	72614440	28.848	27.692



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ050924\
Data File : PQ066448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2024 20:28
Operator : YP\AJ
Sample : P2403-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:16:48 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ042324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 04:24:21 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene


R.T.: 3.440 min
Delta R.T.: 0.000 min
Response: 94169338
Conc: 21.66 ng/ml

Instrument :

ECD_Q


ClientSampleId :

LOD-MDL-WATER-01-QT2-2024




#1 Tetrachloro-m-xylene

R.T.: 2.733 min
Delta R.T.: 0.000 min
Response: 140633086
Conc: 21.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.649 min
Delta R.T.: 0.002 min
Response: 62312200
Conc: 25.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.497 min
Delta R.T.: -0.001 min
Response: 167112699
Conc: 26.56 ng/ml



#3 AR-1016-1


R.T.: 4.567 min
Delta R.T.: 0.017 min
Response: 469584
Conc: 3.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#3 AR-1016-1

R.T.: 3.732 min
Delta R.T.: 0.001 min
Response: 135308
Conc: 0.58 ng/ml




#4 AR-1016-2

R.T.: 4.567 min
Delta R.T.: -0.002 min
Response: 469584
Conc: 2.34 ng/ml



#4 AR-1016-2


R.T.: 3.747 min
Delta R.T.: 0.000 min
Response: 282377
Conc: 0.83 ng/ml



#5 AR-1016-3


R.T.: 4.626 min
Delta R.T.: -0.002 min
Response: 242732
Conc: 1.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#5 AR-1016-3

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 147162
Conc: 0.79 ng/ml




#6 AR-1016-4

R.T.: 4.726 min
Delta R.T.: 0.006 min
Response: 188626
Conc: 1.89 ng/ml



#6 AR-1016-4


R.T.: 3.957 min
Delta R.T.: 0.001 min
Response: 476143
Conc: 3.02 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 5.009 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#7 AR-1016-5

R.T.: 4.149 min
Delta R.T.: 0.000 min
Response: 254393
Conc: 1.31 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 3.642 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 2.931 min
Delta R.T.: -0.004 min
Response: 255424
Conc: 3.03 ng/ml



#9 AR-1221-2


R.T.: 3.721 min
Delta R.T.: 0.000 min
Response: 1320996
Conc: 34.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#9 AR-1221-2

R.T.: 3.000 min
Delta R.T.: -0.011 min
Response: 1961475
Conc: 32.43 ng/ml




#10 AR-1221-3

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 190743
Conc: 1.56 ng/ml



#10 AR-1221-3


R.T.: 3.076 min
Delta R.T.: -0.002 min
Response: 212290
Conc: 1.13 ng/ml



#11 AR-1232-1


R.T.: 3.796 min
Delta R.T.: 0.004 min
Response: 190743
Conc: 1.89 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#11 AR-1232-1

R.T.: 3.076 min
Delta R.T.: -0.002 min
Response: 212290
Conc: 1.38 ng/ml




#12 AR-1232-2

R.T.: 4.279 min
Delta R.T.: -0.011 min
Response: 770317
Conc: 13.97 ng/ml



#12 AR-1232-2


R.T.: 3.747 min
Delta R.T.: 0.000 min
Response: 282377
Conc: 1.75 ng/ml



#13 AR-1232-3


R.T.: 4.567 min
Delta R.T.: -0.003 min
Response: 469584
Conc: 4.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#13 AR-1232-3

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 147162
Conc: 1.67 ng/ml




#14 AR-1232-4

R.T.: 4.726 min
Delta R.T.: 0.006 min
Response: 188626
Conc: 3.98 ng/ml



#14 AR-1232-4


R.T.: 3.988 min
Delta R.T.: -0.004 min
Response: 705366
Conc: 9.36 ng/ml



#15 AR-1232-5


R.T.: 4.823 min
Delta R.T.: 0.006 min
Response: 437928
Conc: 13.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#15 AR-1232-5

R.T.: 4.149 min
Delta R.T.: -0.001 min
Response: 254393
Conc: 3.02 ng/ml




#16 AR-1242-1

R.T.: 4.567 min
Delta R.T.: 0.017 min
Response: 469584
Conc: 4.24 ng/ml



#16 AR-1242-1


R.T.: 3.732 min
Delta R.T.: 0.000 min
Response: 135308
Conc: 0.68 ng/ml



#17 AR-1242-2


R.T.: 4.567 min
Delta R.T.: -0.003 min
Response: 469584
Conc: 2.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#17 AR-1242-2

R.T.: 3.747 min
Delta R.T.: 0.000 min
Response: 282377
Conc: 1.00 ng/ml




#18 AR-1242-3

R.T.: 4.626 min
Delta R.T.: -0.002 min
Response: 242732
Conc: 2.29 ng/ml



#18 AR-1242-3


R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 147162
Conc: 0.94 ng/ml



#19 AR-1242-4


R.T.: 4.726 min
Delta R.T.: 0.006 min
Response: 188626
Conc: 2.26 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#19 AR-1242-4

R.T.: 3.988 min
Delta R.T.: -0.004 min
Response: 705366
Conc: 4.81 ng/ml




#20 AR-1242-5

R.T.: 5.450 min
Delta R.T.: 0.011 min
Response: 265623
Conc: 3.01 ng/ml



#20 AR-1242-5


R.T.: 4.485 min
Delta R.T.: -0.003 min
Response: 654157
Conc: 3.40 ng/ml



#21 AR-1248-1


R.T.: 4.567 min
Delta R.T.: 0.017 min
Response: 469584
Conc: 5.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#21 AR-1248-1

R.T.: 3.732 min
Delta R.T.: 0.000 min
Response: 135308
Conc: 0.91 ng/ml




#22 AR-1248-2

R.T.: 4.823 min
Delta R.T.: 0.006 min
Response: 437928
Conc: 3.80 ng/ml



#22 AR-1248-2


R.T.: 3.957 min
Delta R.T.: 0.001 min
Response: 476143
Conc: 2.15 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 5.009 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#23 AR-1248-3

R.T.: 3.988 min
Delta R.T.: -0.004 min
Response: 705366
Conc: 3.34 ng/ml




#24 AR-1248-4

R.T.: 5.402 min
Delta R.T.: -0.002 min
Response: 2763472
Conc: 19.18 ng/ml



#24 AR-1248-4


R.T.: 4.149 min
Delta R.T.: 0.000 min
Response: 254393
Conc: 0.99 ng/ml



#25 AR-1248-5


R.T.: 5.450 min
Delta R.T.: 0.010 min
Response: 265623
Conc: 1.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#25 AR-1248-5

R.T.: 4.509 min
Delta R.T.: -0.014 min
Response: 687344
Conc: 2.66 ng/ml




#26 AR-1254-1

R.T.: 5.402 min
Delta R.T.: 0.026 min
Response: 2763472
Conc: 18.39 ng/ml



#26 AR-1254-1


R.T.: 4.485 min
Delta R.T.: 0.000 min
Response: 654157
Conc: 1.66 ng/ml



#27 AR-1254-2


R.T.: 5.592 min
Delta R.T.: -0.001 min
Response: 364198
Conc: 1.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#27 AR-1254-2

R.T.: 4.628 min
Delta R.T.: -0.003 min
Response: 606215
Conc: 1.78 ng/ml




#28 AR-1254-3

R.T.: 5.952 min
Delta R.T.: 0.003 min
Response: 212511
Conc: 0.86 ng/ml



#28 AR-1254-3


R.T.: 5.027 min
Delta R.T.: 0.014 min
Response: 4976660
Conc: 9.21 ng/ml



#29 AR-1254-4


R.T.: 6.248 min
Delta R.T.: 0.013 min
Response: 1517436
Conc: 8.71 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#29 AR-1254-4

R.T.: 5.246 min
Delta R.T.: 0.006 min
Response: 709861
Conc: 1.91 ng/ml




#30 AR-1254-5

R.T.: 6.638 min
Delta R.T.: -0.011 min
Response: 302838
Conc: 1.65 ng/ml



#30 AR-1254-5


R.T.: 5.646 min
Delta R.T.: 0.002 min
Response: 296006
Conc: 0.61 ng/ml



#31 AR-1260-1


R.T.: 6.109 min
Delta R.T.: -0.006 min
Response: 500294
Conc: 2.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#31 AR-1260-1

R.T.: 5.142 min
Delta R.T.: 0.000 min
Response: 647041
Conc: 1.65 ng/ml




#32 AR-1260-2

R.T.: 6.418 min
Delta R.T.: 0.043 min
Response: 246966
Conc: 1.17 ng/ml



#32 AR-1260-2


R.T.: 5.327 min
Delta R.T.: -0.005 min
Response: 1472830
Conc: 3.09 ng/ml



#33 AR-1260-3


R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 2491259
Conc: 16.20 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#33 AR-1260-3

R.T.: 5.473 min
Delta R.T.: -0.001 min
Response: 1069616
Conc: 2.43 ng/ml




#34 AR-1260-4

R.T.: 6.970 min
Delta R.T.: 0.023 min
Response: 3742460
Conc: 21.52 ng/ml



#34 AR-1260-4


R.T.: 5.933 min
Delta R.T.: -0.002 min
Response: 8201868
Conc: 21.54 ng/ml



#35 AR-1260-5


R.T.: 7.259 min
Delta R.T.: 0.000 min
Response: 1211875
Conc: 3.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#35 AR-1260-5

R.T.: 6.181 min
Delta R.T.: -0.001 min
Response: 3776369
Conc: 4.61 ng/ml




#36 AR-1262-1

R.T.: 6.638 min
Delta R.T.: -0.009 min
Response: 302838
Conc: 2.06 ng/ml



#36 AR-1262-1


R.T.: 5.646 min
Delta R.T.: 0.001 min
Response: 296006
Conc: 1.06 ng/ml



#37 AR-1262-2


R.T.: 7.259 min
Delta R.T.: 0.000 min
Response: 1211875
Conc: 3.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#37 AR-1262-2

R.T.: 5.933 min
Delta R.T.: -0.003 min
Response: 8201868
Conc: 17.25 ng/ml




#38 AR-1262-3

R.T.: 7.534 min
Delta R.T.: -0.002 min
Response: 12786872
Conc: 53.14 ng/ml



#38 AR-1262-3


R.T.: 6.458 min
Delta R.T.: -0.002 min
Response: 33280187
Conc: 88.12 ng/ml



#39 AR-1262-4


R.T.: 7.612 min
Delta R.T.: 0.002 min
Response: 11810023
Conc: 63.61 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#39 AR-1262-4

R.T.: 6.521 min
Delta R.T.: 0.000 min
Response: 28727886
Conc: 41.16 ng/ml




#40 AR-1262-5

R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 4097951
Conc: 33.13 ng/ml



#40 AR-1262-5


R.T.: 7.014 min
Delta R.T.: -0.001 min
Response: 10072684
Conc: 30.61 ng/ml



#41 AR-1268-1


R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 12786872
Conc: 30.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#41 AR-1268-1

R.T.: 6.458 min
Delta R.T.: -0.003 min
Response: 33280187
Conc: 29.70 ng/ml




#42 AR-1268-2

R.T.: 7.612 min
Delta R.T.: -0.001 min
Response: 11810023
Conc: 30.36 ng/ml



#42 AR-1268-2


R.T.: 6.521 min
Delta R.T.: -0.003 min
Response: 28727886
Conc: 28.16 ng/ml



#43 AR-1268-3


R.T.: 7.793 min
Delta R.T.: 0.000 min
Response: 9937744
Conc: 30.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024




#43 AR-1268-3

R.T.: 6.720 min
Delta R.T.: -0.003 min
Response: 25338310
Conc: 28.49 ng/ml




#44 AR-1268-4

R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 4097951
Conc: 28.92 ng/ml



#44 AR-1268-4


R.T.: 7.014 min
Delta R.T.: -0.003 min
Response: 10072684
Conc: 27.08 ng/ml



#45 AR-1268-5

R.T.: 8.413 min
Delta R.T.: 0.000 min
Response: 27342077
Conc: 28.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024



#45 AR-1268-5

R.T.: 7.282 min
Delta R.T.: -0.003 min
Response: 72614440
Conc: 27.69 ng/ml