

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051022\
 Data File : PQ057554.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2022 15:42
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543236

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 03:08:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 06 23:16:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.521	3.739	646.0E6	269.8E6	18.905	20.286
2) SA Decachlor...	10.279	8.659	1195.9E6	471.2E6	40.214	43.320
Target Compounds						
26) L6 AR-1254-1	6.542	5.586	319.5E6	295.0E6	358.794	387.738
27) L6 AR-1254-2	6.760	5.731	507.6E6	240.3E6	361.711	390.669
28) L6 AR-1254-3	7.131	6.129	596.9E6	395.4E6	356.309	382.643
29) L6 AR-1254-4	7.420	6.358	390.1E6	284.2E6	354.304	374.251
30) L6 AR-1254-5	7.833	6.764	525.3E6	352.6E6	421.650	377.946

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051022\
Data File : PQ057554.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2022 15:42
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543236

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 03:08:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 06 23:16:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

