

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051122\
 Data File : PQ057576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2022 11:23
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12543244

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 01:42:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 06 23:16:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.521	3.739	644.2E6	261.2E6	18.852	19.638
2) SA Decachlor...	10.278	8.658	1169.9E6	441.1E6	39.338	40.549
Target Compounds						
26) L6 AR-1254-1	6.542	5.585	314.5E6	282.3E6	353.112	370.998
27) L6 AR-1254-2	6.760	5.730	498.8E6	229.0E6	355.467	372.242
28) L6 AR-1254-3	7.131	6.128	584.7E6	376.0E6	349.035	363.953
29) L6 AR-1254-4	7.420	6.357	383.9E6	272.1E6	348.639	358.314
30) L6 AR-1254-5	7.834	6.763	434.3E6	334.1E6	348.610	358.118

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051122\
Data File : PQ057576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2022 11:23
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12543244

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 01:42:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 06 23:16:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

