

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051419\
 Data File : PQ039750.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2019 21:57
 Operator : AJ\MA
 Sample : K2768-19
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P026-SS016-0612-01

Manual Integrations
 APPROVED

Ankita
 5/14/2019 11:49:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 04:58:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 10 01:55:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.643	4.751	128.2E6	90492232	22.431m	24.658m
2) SA Decachlor...	12.010	10.498	88545792	52272859	13.704m	15.776m
Target Compounds						
31) L7 AR-1260-1	8.859	7.911	8586203	5502963	33.674m	31.073m
32) L7 AR-1260-2	9.131	8.116	20132139	17168754	64.592m	78.246m
33) L7 AR-1260-3	9.506	8.279	13656159	10098623	54.633m	48.831m
34) L7 AR-1260-4	9.747	8.777	11016691	6807116	38.184m	39.012m
35) L7 AR-1260-5	10.092	9.030	23963701	16666759	40.543m	39.312m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051419\
Data File : PQ039750.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2019 21:57
Operator : AJ\MA
Sample : K2768-19
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
P026-SS016-0612-01


Manual Integrations
APPROVED

Ankita
5/14/2019 11:49:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 04:58:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050919.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 10 01:55:02 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.643 min
Delta R.T.: -0.023 min
Response: 128223130
Conc: 22.43 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
P026-SS016-0612-01


Manual Integrations
APPROVED

Ankita
5/14/2019 11:49:50 AM




#1 Tetrachloro-m-xylene

R.T.: 4.751 min
Delta R.T.: -0.015 min
Response: 90492232
Conc: 24.66 ng/ml m




#2 Decachlorobiphenyl

R.T.: 12.010 min
Delta R.T.: -0.039 min
Response: 88545792
Conc: 13.70 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.498 min
Delta R.T.: -0.023 min
Response: 52272859
Conc: 15.78 ng/ml m




#31 AR-1260-1

R.T.: 8.859 min
Delta R.T.: -0.023 min
Response: 8586203
Conc: 33.67 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
P026-SS016-0612-01


Manual Integrations
APPROVED

Ankita
5/14/2019 11:49:50 AM




#31 AR-1260-1

R.T.: 7.911 min
Delta R.T.: -0.016 min
Response: 5502963
Conc: 31.07 ng/ml m




#32 AR-1260-2

R.T.: 9.131 min
Delta R.T.: -0.023 min
Response: 20132139
Conc: 64.59 ng/ml m



#32 AR-1260-2

R.T.: 8.116 min
Delta R.T.: -0.016 min
Response: 17168754
Conc: 78.25 ng/ml m




#33 AR-1260-3

R.T.: 9.506 min
Delta R.T.: -0.022 min
Response: 13656159
Conc: 54.63 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
P026-SS016-0612-01


Manual Integrations
APPROVED

Ankita
5/14/2019 11:49:50 AM




#33 AR-1260-3

R.T.: 8.279 min
Delta R.T.: -0.018 min
Response: 10098623
Conc: 48.83 ng/ml m




#34 AR-1260-4

R.T.: 9.747 min
Delta R.T.: -0.026 min
Response: 11016691
Conc: 38.18 ng/ml m



#34 AR-1260-4

R.T.: 8.777 min
Delta R.T.: -0.017 min
Response: 6807116
Conc: 39.01 ng/ml m




#35 AR-1260-5

R.T.: 10.092 min
Delta R.T.: -0.024 min
Response: 23963701
Conc: 40.54 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
P026-SS016-0612-01

Manual Integrations
APPROVED

Ankita
5/14/2019 11:49:50 AM



#35 AR-1260-5

R.T.: 9.030 min
Delta R.T.: -0.017 min
Response: 16666759
Conc: 39.31 ng/ml m