

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051619.M
 Title : GC EXTRACTABLES
 Last Update : Thu May 16 02:01:43 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039781.D 1000 =PQ039779.D 750 =PQ039780.D
 250 =PQ039782.D 50 =PQ039783.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.104	4.755	4.919	4.832	5.354	4.993 E6	4.81	
2) SA	Decachlorobiphenyl	5.177	5.060	5.070	5.260	6.054	5.324 E6	7.81	
3) L1	AR-1016-1	1.697	1.532	1.633	1.687	2.024	1.715 E5	10.78	
4) L1	AR-1016-2	2.495	2.300	2.450	2.518	2.925	2.538 E5	9.17	
5) L1	AR-1016-3	1.519	1.395	1.506	1.581	1.841	1.568 E5	10.64	
6) L1	AR-1016-4	1.255	1.146	1.240	1.258	1.458	1.271 E5	8.98	
7) L1	AR-1016-5	1.219	1.119	1.190	1.210	1.307	1.209 E5	5.59	
8) L2	AR-1221-1	5.277					5.277 E4	0.00	
9) L2	AR-1221-2	4.205					4.205 E4	0.00	
10) L2	AR-1221-3	1.334					1.334 E5	0.00	
11) L3	AR-1232-1	1.374					1.374 E5	0.00	
12) L3	AR-1232-2	7.306					7.306 E4	0.00	
13) L3	AR-1232-3	1.486					1.486 E5	0.00	
14) L3	AR-1232-4	7.423					7.423 E4	0.00	
15) L3	AR-1232-5	4.920					4.920 E4	0.00	
16) L4	AR-1242-1	1.313					1.313 E5	0.00	
17) L4	AR-1242-2	1.925					1.925 E5	0.00	
18) L4	AR-1242-3	1.205					1.205 E5	0.00	
19) L4	AR-1242-4	9.759					9.759 E4	0.00	
20) L4	AR-1242-5	1.098					1.098 E5	0.00	
21) L5	AR-1248-1	1.019					1.019 E5	0.00	
22) L5	AR-1248-2	1.326					1.326 E5	0.00	
23) L5	AR-1248-3	1.596					1.596 E5	0.00	
24) L5	AR-1248-4	1.800					1.800 E5	0.00	
25) L5	AR-1248-5	1.769					1.769 E5	0.00	
26) L6	AR-1254-1	2.207					2.207 E5	0.00	
27) L6	AR-1254-2	3.429					3.429 E5	0.00	
28) L6	AR-1254-3	3.680					3.680 E5	0.00	
29) L6	AR-1254-4	2.699					2.699 E5	0.00	
30) L6	AR-1254-5	2.920					2.920 E5	0.00	
31) L7	AR-1260-1	2.177	2.040	2.135	2.189	2.486	2.205 E5	7.58	
32) L7	AR-1260-2	2.668	2.472	2.586	2.626	3.069	2.684 E5	8.46	
33) L7	AR-1260-3	2.160	1.987	2.098	2.102	2.585	2.186 E5	10.58	
34) L7	AR-1260-4	2.468	2.282	2.370	2.389	3.098	2.521 E5	13.06	
35) L7	AR-1260-5	4.926	4.801	4.887	4.806	5.607	5.005 E5	6.80	
36) L8	AR-1262-1	3.289					3.289 E5	0.00	
37) L8	AR-1262-2	5.839					5.839 E5	0.00	
38) L8	AR-1262-3	3.940					3.940 E5	0.00	
39) L8	AR-1262-4	3.136					3.136 E5	0.00	
40) L8	AR-1262-5	2.006					2.006 E5	0.00	
41) L9	AR-1268-1	6.408					6.408 E5	0.00	
42) L9	AR-1268-2	6.309					6.309 E5	0.00	
43) L9	AR-1268-3	5.461					5.461 E5	0.00	
44) L9	AR-1268-4	2.204					2.204 E5	0.00	
45) L9	AR-1268-5	1.867					1.867 E6	0.00	

Signal #2 Calibration Files

500 =PQ039781.D 1000 =PQ039779.D 750 =PQ039780.D
 250 =PQ039782.D 50 =PQ039783.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.060	3.866	4.005	3.824	4.068	3.964 E6	2.84	
2) SA	Decachlorobiphenyl	3.559	3.512	3.479	3.511	4.335	3.679 E6	9.99	
3) L1	AR-1016-1	1.464	1.374	1.449	1.452	1.650	1.478 E5	6.94	
4) L1	AR-1016-2	2.104	1.971	2.070	2.058	2.437	2.128 E5	8.45	
5) L1	AR-1016-3	1.143	1.071	1.108	1.127	1.594	1.209 E5	17.99	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051619.M
 Title : GC EXTRACTABLES
 Last Update : Thu May 16 02:01:43 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ039781.D 1000 =PQ039779.D 750 =PQ039780.D
 250 =PQ039782.D 50 =PQ039783.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	0.882	0.825	0.861	0.865	1.223	0.931	E5	17.65
7)	L1 AR-1016-5	1.122	1.091	1.098	1.175	1.528	1.203	E5	15.35
8)	L2 AR-1221-1	4.166					4.166	E4	0.00
9)	L2 AR-1221-2	3.212					3.212	E4	0.00
10)	L2 AR-1221-3	1.061					1.061	E5	0.00
11)	L3 AR-1232-1	1.076					1.076	E5	0.00
12)	L3 AR-1232-2	1.242					1.242	E5	0.00
13)	L3 AR-1232-3	6.742					6.742	E4	0.00
14)	L3 AR-1232-4	5.534					5.534	E4	0.00
15)	L3 AR-1232-5	6.611					6.611	E4	0.00
16)	L4 AR-1242-1	1.126					1.126	E5	0.00
17)	L4 AR-1242-2	1.591					1.591	E5	0.00
18)	L4 AR-1242-3	8.096					8.096	E4	0.00
19)	L4 AR-1242-4	7.849					7.849	E4	0.00
20)	L4 AR-1242-5	1.137					1.137	E5	0.00
21)	L5 AR-1248-1	8.664					8.664	E4	0.00
22)	L5 AR-1248-2	1.132					1.132	E5	0.00
23)	L5 AR-1248-3	1.173					1.173	E5	0.00
24)	L5 AR-1248-4	1.495					1.495	E5	0.00
25)	L5 AR-1248-5	1.490					1.490	E5	0.00
26)	L6 AR-1254-1	2.792					2.792	E5	0.00
27)	L6 AR-1254-2	2.457					2.457	E5	0.00
28)	L6 AR-1254-3	3.956					3.956	E5	0.00
29)	L6 AR-1254-4	2.488					2.488	E5	0.00
30)	L6 AR-1254-5	3.381					3.381	E5	0.00
31)	L7 AR-1260-1	2.093	1.941	2.040	2.050	2.359	2.097	E5	7.49
32)	L7 AR-1260-2	2.590	2.438	2.541	2.588	3.032	2.638	E5	8.68
33)	L7 AR-1260-3	2.366	2.241	2.338	2.303	2.802	2.410	E5	9.29
34)	L7 AR-1260-4	1.969	1.877	1.947	1.923	2.269	1.997	E5	7.80
35)	L7 AR-1260-5	4.867	4.753	4.863	4.708	5.103	4.859	E5	3.15
36)	L8 AR-1262-1	1.556					1.556	E5	0.00
37)	L8 AR-1262-2	5.611					5.611	E5	0.00
38)	L8 AR-1262-3	2.203					2.203	E5	0.00
39)	L8 AR-1262-4	3.974					3.974	E5	0.00
40)	L8 AR-1262-5	1.595					1.595	E5	0.00
41)	L9 AR-1268-1	5.473					5.473	E5	0.00
42)	L9 AR-1268-2	5.202					5.202	E5	0.00
43)	L9 AR-1268-3	4.290					4.290	E5	0.00
44)	L9 AR-1268-4	1.720					1.720	E5	0.00
45)	L9 AR-1268-5	1.398					1.398	E6	0.00

(#) = Out of Range