

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051819CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri May 17 18:06:14 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ039862.D 1600 =PQ039864.D 800 =PQ039863.D
 200 =PQ039861.D 100 =PQ039860.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	7.042	7.897	7.116	7.099	7.673	7.366	E6	5.32
2) SA	Decachlorobiphenyl	6.433	5.052	5.917	6.940	7.042	6.277	E6	13.04
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	1.250	1.265	1.207	1.312	1.459	1.299	E5	7.48
22) L5	AR-1248-2	1.594	1.590	1.527	1.703	1.795	1.642	E5	6.47
23) L5	AR-1248-3	1.885	1.911	1.827	1.996	2.086	1.941	E5	5.22
24) L5	AR-1248-4	2.092	2.107	2.015	2.222	2.355	2.158	E5	6.14
25) L5	AR-1248-5	2.007	2.058	1.968	2.153	2.237	2.085	E5	5.26
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ039862.D 1600 =PQ039864.D 800 =PQ039863.D
 200 =PQ039861.D 100 =PQ039860.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.277	6.056	5.447	5.314	5.670	5.553	E6	5.77
2) SA	Decachlorobiphenyl	4.153	3.478	3.931	4.349	4.387	4.060	E6	9.17
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051819CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri May 17 18:06:14 2019
 Response Via : Initial Calibration

Calibration Files

400 =PQ039862.D 1600 =PQ039864.D 800 =PQ039863.D
 200 =PQ039861.D 100 =PQ039860.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	0.986	1.041	0.968	1.007	1.050	1.010 E5	3.46
22) L5	AR-1248-2	1.248	1.301	1.222	1.255	1.305	1.266 E5	2.84
23) L5	AR-1248-3	1.305	1.345	1.266	1.311	1.405	1.326 E5	3.93
24) L5	AR-1248-4	1.595	1.671	1.565	1.615	1.746	1.638 E5	4.38
25) L5	AR-1248-5	1.571	1.641	1.541	1.590	1.759	1.620 E5	5.27
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range