Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051919\
Data File : PQ040006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2019 14:40
Operator : SM\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248325

Manual Integrations APPROVED
Ankita
5/20/2019 2:00:51 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 05:14:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 18 00:58:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.608	4.730	167.2E6	121.7E6	24.793	25.329
2)	SA Decachlor...	11.944	10.466	198.2E6	123.4E6	31.735	31.972
	Target Compounds						
21)	L5 AR-1248-1	7.060	6.181	49179214	40934949	378.697m	405.100m
22)	L5 AR-1248-2	7.373	6.472	64375325	52359157	392.092m	413.493
23)	L5 AR-1248-3	7.603	6.519	73538013	53030058	378.877	399.823
24)	L5 AR-1248-4	8.051	6.724	68673001	64032331	318.191	390.839m
25)	L5 AR-1248-5	8.092	7.180	66956366	50534248	321.208	311.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051919\
Data File : PQ040006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2019 14:40
Operator : SM\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248325

Manual Integrations APPROVED
Ankita
5/20/2019 2:00:51 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 05:14:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 18 00:58:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm