

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052019\
 Data File : PQ040037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2019 19:14
 Operator : SM\AJ
 Sample : AR1268CCC400
 Misc :
 ALS Vial : 145 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268327

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 00:47:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 18 00:58:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.626	4.751	101.5E6	75977382	15.052	15.819
2) SA Decachlor...	11.966	10.483	602.3E6	370.6E6	96.445	96.000
Target Compounds						
41) L9 AR-1268-1	10.391	9.312	542.6E6	424.4E6	448.056	410.194
42) L9 AR-1268-2	10.492	9.380	519.5E6	389.9E6	451.356	407.436
43) L9 AR-1268-3	10.728	9.588	483.0E6	346.5E6	476.085	430.854
44) L9 AR-1268-4	11.186	9.903	159.5E6	116.0E6	422.170	386.682
45) L9 AR-1268-5	11.617	10.214	1549.5E6	1030.9E6	467.606	419.330

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052019\
Data File : PQ040037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2019 19:14
Operator : SM\AJ
Sample : AR1268CCC400
Misc :
ALS Vial : 145 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268327

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 00:47:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 18 00:58:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

