Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052019\
Data File : PQ040062.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2019 06:51
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254329

Manual Integrations APPROVED
Ankita
5/21/2019 9:53:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 07:04:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 18 00:58:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.626	4.752	84743845	61271514	12.570	12.757
2)	SA Decachlor...	11.955	10.477	234.4E6	137.9E6	37.539	35.711
	Target Compounds						
26)	L6 AR-1254-1	8.031	7.147	66453385	76561652	287.871	289.077
27)	L6 AR-1254-2	8.268	7.314	103.3E6	68322795	286.611	299.289
28)	L6 AR-1254-3	8.664	7.761	108.0E6	110.3E6	284.447	298.434
29)	L6 AR-1254-4	8.965	8.011	65634128	56946805	240.473	248.116m
30)	L6 AR-1254-5	9.402	8.455	84654723	90300120	281.694m	290.035

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052019\
Data File : PQ040062.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2019 06:51
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254329

Manual Integrations APPROVED
Ankita
5/21/2019 9:53:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 07:04:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 18 00:58:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm