

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052021\
 Data File : PQ053631.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2021 23:35
 Operator : DD\AJ
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 04:33:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 04:32:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.217	4.232	886.9E6	441.5E6	38.640	39.451
2)	SA Decachlor...	11.401	9.568	1893.2E6	1043.1E6	77.077	78.117
Target Compounds							
21)	L5 AR-1248-1	6.542	5.491	384.5E6	198.6E6	768.453	753.870
22)	L5 AR-1248-2	6.835	5.754	571.0E6	280.9E6	759.889	748.770
23)	L5 AR-1248-3	7.054	5.799	658.2E6	291.5E6	769.959	751.625
24)	L5 AR-1248-4	7.483	5.984	801.7E6	360.3E6	769.339	758.109
25)	L5 AR-1248-5	7.524	6.407	787.5E6	387.6E6	760.255	760.271



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052021\
Data File : PQ053631.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2021 23:35
Operator : DD\AJ
Sample : AR1248ICC800
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 04:33:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 20 04:32:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 886856716
Conc: 38.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.232 min
Delta R.T.: 0.000 min
Response: 441544032
Conc: 39.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.401 min
Delta R.T.: 0.001 min
Response: 1893225667
Conc: 77.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.568 min
Delta R.T.: 0.000 min
Response: 1043075732
Conc: 78.12 ng/ml



#21 AR-1248-1


R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 384507147
Conc: 768.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#21 AR-1248-1

R.T.: 5.491 min
Delta R.T.: 0.000 min
Response: 198633877
Conc: 753.87 ng/ml




#22 AR-1248-2

R.T.: 6.835 min
Delta R.T.: 0.000 min
Response: 570970364
Conc: 759.89 ng/ml



#22 AR-1248-2


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 280918468
Conc: 748.77 ng/ml



#23 AR-1248-3


R.T.: 7.054 min
Delta R.T.: 0.000 min
Response: 658220852
Conc: 769.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#23 AR-1248-3

R.T.: 5.799 min
Delta R.T.: 0.000 min
Response: 291475933
Conc: 751.63 ng/ml




#24 AR-1248-4

R.T.: 7.483 min
Delta R.T.: 0.000 min
Response: 801691267
Conc: 769.34 ng/ml



#24 AR-1248-4


R.T.: 5.984 min
Delta R.T.: 0.000 min
Response: 360339462
Conc: 758.11 ng/ml



#25 AR-1248-5

R.T.: 7.524 min
Delta R.T.: 0.000 min
Response: 787501565
Conc: 760.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#25 AR-1248-5

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 387631842
Conc: 760.27 ng/ml