

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052021\
 Data File : PQ053632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2021 23:51
 Operator : DD\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 04:34:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 04:32:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.217	4.232	1733.9E6	885.9E6	75.543	79.157
2) SA Decachlor...	11.402	9.568	3631.4E6	2008.2E6	147.843	150.392
Target Compounds						
21) L5 AR-1248-1	6.542	5.491	749.7E6	388.9E6	1498.374	1475.953
22) L5 AR-1248-2	6.834	5.754	1086.0E6	547.4E6	1445.264	1459.005
23) L5 AR-1248-3	7.054	5.799	1257.9E6	568.5E6	1471.471	1466.052
24) L5 AR-1248-4	7.483	5.984	1523.2E6	708.7E6	1461.694	1490.926
25) L5 AR-1248-5	7.523	6.407	1495.4E6	762.0E6	1443.683	1494.462



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052021\
Data File : PQ053632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2021 23:51
Operator : DD\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 04:34:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 20 04:32:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 1733856891
Conc: 75.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.232 min
Delta R.T.: 0.000 min
Response: 885939569
Conc: 79.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.402 min
Delta R.T.: 0.002 min
Response: 3631424145
Conc: 147.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.568 min
Delta R.T.: 0.000 min
Response: 2008161104
Conc: 150.39 ng/ml



#21 AR-1248-1


R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 749734190
Conc: 1498.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#21 AR-1248-1

R.T.: 5.491 min
Delta R.T.: 0.000 min
Response: 388892425
Conc: 1475.95 ng/ml




#22 AR-1248-2

R.T.: 6.834 min
Delta R.T.: 0.000 min
Response: 1085951926
Conc: 1445.26 ng/ml



#22 AR-1248-2


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 547379354
Conc: 1459.00 ng/ml



#23 AR-1248-3


R.T.: 7.054 min
Delta R.T.: 0.000 min
Response: 1257927724
Conc: 1471.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#23 AR-1248-3

R.T.: 5.799 min
Delta R.T.: 0.000 min
Response: 568526324
Conc: 1466.05 ng/ml




#24 AR-1248-4

R.T.: 7.483 min
Delta R.T.: 0.000 min
Response: 1523160191
Conc: 1461.69 ng/ml



#24 AR-1248-4


R.T.: 5.984 min
Delta R.T.: 0.000 min
Response: 708657251
Conc: 1490.93 ng/ml



#25 AR-1248-5

R.T.: 7.523 min
Delta R.T.: 0.000 min
Response: 1495423389
Conc: 1443.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#25 AR-1248-5

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 761966504
Conc: 1494.46 ng/ml