

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052219\
Data File : PQ040077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2019 12:42
Operator : SM\AJ
Sample : AR1232ICC200
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1232201

Manual Integrations
APPROVED

Ankita
5/23/2019 10:32:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:15:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 14:14:36 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.612min 11.330 ng/ml

response 36720329

(1) Tetrachloro-m-xylene #2 (SA)

4.741min 11.906 ng/ml

response 25335563

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052219\
Data File : PQ040077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2019 12:42
Operator : SM\AJ
Sample : AR1232ICC200
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1232201

Manual Integrations
APPROVED

Ankita
5/23/2019 10:32:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:15:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 14:14:36 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.612min 11.330 ng/ml

response 36720329

(1) Tetrachloro-m-xylene #2 (SA)

4.740min 10.904 ng/ml m

response 23203927

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052219\
 Data File : PQ040077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2019 12:42
 Operator : SM\AJ
 Sample : AR1232ICC200
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1232201

Manual Integrations
 APPROVED

Ankita
 5/23/2019 10:32:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 14:15:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 22 14:14:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.612	4.740	36720329	23203927	11.330	10.904m
2) SA Decachlor...	11.934	10.467	147.5E6	83883036	21.390	20.755
Target Compounds						
11) L3 AR-1232-1	6.100	5.264	14044891	10189335	226.353	226.876
12) L3 AR-1232-2	6.744	6.208	8716482	14258537	224.307	243.557
13) L3 AR-1232-3	7.084	6.423	17813033	8421684	226.070	243.948
14) L3 AR-1232-4	7.270	6.525	9200906	6318898	210.357	212.763
15) L3 AR-1232-5	7.373	6.728	5863240	6139115	196.881	181.888

17
 05/29/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.