

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052521\
Data File : PQ053660.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2021 20:20
Operator : DD\AJ
Sample : M2511-02
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleID :
C0CL3

Manual Integrations
APPROVED

Ankita
5/27/2021 11:32:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 06:04:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 20 06:23:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
7.45 5289805485 3721.60
7.68 9045391222 4020.83
8.06 10797534864 4556.10
8.36 8980789643 4876.67
8.79 15072214808 8340.69
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.36 3556045290 3144.53
6.52 4089331558 4019.77
6.94 7187407519 4294.75
7.18 5504215130 4813.51
7.61 10818933718 6678.14

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052521\
 Data File : PQ053660.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2021 20:20
 Operator : DD\AJ
 Sample : M2511-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1


Instrument :
 ECD_Q
 ClientSampleID :
 C0CL3

Manual Integrations
 APPROVED

Ankita
 5/27/2021 11:32:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 06:04:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 06:23:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 7.45 5289805485 3721.60
 7.68 9045391222 4020.83
 8.06 10797534864 4556.10
 8.36 8980789643 4876.67
 8.79 10433059507 5773.47

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 6.36 3556045290 3144.53
 6.52 4089331558 4019.77
 6.94 7187407519 4294.75
 7.18 5504215130 4813.51
 7.61 10818933718 6678.14

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052521\
 Data File : PQ053660.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2021 20:20
 Operator : DD\AJ
 Sample : M2511-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 C0CL3

Manual Integrations
 APPROVED

Ankita
 5/27/2021 11:32:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 06:04:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 06:23:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	5.218	4.234	450.5E6	231.9E6	19.866	21.531
2) SA Decachlor...	11.400	9.568	933.8E6	503.0E6	41.484	45.548

Target Compounds

26) L6 AR-1254-1	7.452	6.362	5289.8E6	3556.0E6	3721.602	3144.533
27) L6 AR-1254-2	7.679	6.521	9045.4E6	4089.3E6	4020.825	4019.769
28) L6 AR-1254-3	8.063	6.942	10797.5E6	7187.4E6	4556.102	4294.749
29) L6 AR-1254-4	8.358	7.180	8980.8E6	5504.2E6	4876.670	4813.514
30) L6 AR-1254-5	8.787	7.609	10433.1E6	10818.9E6	5773.469m)	6678.139

AJ
 05/28/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.